

From: [Tran, Frank](#)
To: [M. Srinivasan](#)
Subject: Request additional information for NRC License No. 21-00741-08
Date: Thursday, May 27, 2021 1:53:00 PM

Dear Ms. Srinivasan:

We have reviewed the letters dated May 5, 2021 and March 19, 2018, as referenced. We will need the following information.

Item 2 in letter dated March 19, 2018 said the pressure in the glove box fitted with HEPA / charcoal filter where DU will be used is positive. I think it should be negative (lower than the surrounding areas). Could you please help me understand more on this.

Item 2 in letter dated May 5, 2021 said washing and drying DU will be performed in the chemical fume hood. Are the fume hood fitted with filters (if yes, are they HEPA / charcoal filters?) to mitigate the potential airborne or air effluent release?

To continue review of your application, we request that you submit your response with date and authorized signature to this correspondence within 30 calendar days from the date of this email. In your response, please refer the license number, docket number and Mail Control No. 625560.

If you have questions, require additional time to respond, or require clarification on any of the information stated above, please contact me at 630-829-9623 or reply to this email.

In accordance with Title 10 of the Code of Federal Regulations (10 CFR) 2.390 of the NRC's "Rules of Practice," a copy of this correspondence will be made available electronically for public inspection in the NRC Public Document Room or from the NRC's Agencywide Documents Access and Management System (ADAMS), accessible from the NRC Web site at <https://www.nrc.gov/reading-rm/adams.html>.

Best regards,

Frank Tran

Health Physicist/License Reviewer
NRC Region III/Division of Nuclear Materials Safety
Phone: 630-829-9623
Fax: 630-515-1078
Email: Frank.Tran@nrc.gov

