



525 John Street – Second Floor  
Bridgeport, CT 06604  
Ph: 203-336-5900  
Fax: 203-336-1769  
www.fairfieldtesting.com

Briana G. DeBoer  
Nuclear Regulatory Commission  
Region 1

June 18, 2021

SUBJECT: Mail Control No. 625514, request for additional information

Dear Ms. DeBoer,

Please find the additional information requested from your letter date June 15, 2021 pertaining to our license renewal, License No. 06-19279-01.

1)

a. Mr. Kieley is both the RSO and lab/field manager and is authorized to make legally binding commitments on behalf of Megan, LLC.

2)

a. Yes, we would like to be authorized for a maximum activity per Cs-137 source to 9mCi and Am-241 per source to 44 mCi for the Troxler Model 3440 series.

b. Yes, we would like to be authorized for a maximum activity per CS-137 source to 11mCi and Am-241 per source to 44 mCi for the Humboldt Scientific, Inc. Model 5001.

c. We have never possessed devices of the Troxler 4640 series or InstroTek, Inc. Model 3500. From what I can tell. these models were added as an amendment in case we ever had the need to posses them.

d. We would like to keep our possession limit at 110 mCi for Cs-137 and 550 mCi for Am-241 as our total activity with our responses to 2a. and 2b. would still be below these limits.

If you require any other information, feel free to contact myself or Mr. Kieley and we will send any more required information over promptly.

Thank you,

  
James Quill, Principal