



Region I Office
 Division of Nuclear Materials Safety
 2100 Renaissance Boulevard, Suite 100
 King of Prussia, PA 19406-2713
 (610)337-5000

Telephone Conversation Record

Date: 7/21/21

License No. 31-31284-01

Docket No.(no hyphens): 030-37574

Mail Control/Report No. 627315

Licensee Name: Advance Testing Company, Inc.

Participant(s) Name/Title: Patrick Kiernan

Work Telephone No. 845-496-1600, ext 211

Business Cellphone No. N/A

NRC Representative Name/Title: Jonathan Pfungsten, Health Physicist

Subject: Amendment Request

Discussion: Region I received a license amendment request dated June 4, 2021, requesting a change in leak testing requirements from a 6-month frequency to 12 months. However, as discussed on the phone, the license condition states "Sealed sources shall be tested for leakage and/or contamination **at intervals not to exceed the intervals specified in the certificate of registration** issued by the U.S. Nuclear Regulatory Commission under 10 CFR 32.210 or by an Agreement State." In this licensee's case, the certificates of registration for the three authorized device models (i.e. Troxler Electronic Laboratories Models 4640 and 3400 series and InstroTek, Inc. Model 3500 series) all allow for a 12-month leak test frequency already.

The licensee agreed that, in light of this information, the amendment was no longer required. They intend to withdraw the amendment request.

Action Required: Await withdrawal letter, process closure of licensing action

SUNSI REVIEW		
Document Availability:	<input checked="" type="checkbox"/> Public	or <input type="checkbox"/> Non-Public
Document Sensitivity: (select "1" value to the right)	<input checked="" type="checkbox"/> Non-Sensitive	<input type="checkbox"/> MD 3.4 Non-Public B.1 (Non-Sensitive)
	<input type="checkbox"/> Non-Sensitive Copyright	<input type="checkbox"/> MD 3.4 Non-Public A.3 (Sensitive Security Related)
	<input type="checkbox"/> Sensitive – Proprietary	<input type="checkbox"/> MD 3.4 Non-Public A.4 (Sensitive Proprietary)
	<input type="checkbox"/> Sensitive – Privacy Act (includes Personally Identifiable Information (PII))	<input type="checkbox"/> MD 3.4 Non-Public A.5 (Sensitive-Privacy Act)
	<input type="checkbox"/> Sensitive – Internal, Periodic Review required (All Other Sensitive Internal Info.)	<input type="checkbox"/> MD 3.4 Non-Public A.6 (Sensitive-Federal, State, etc)
	<input type="checkbox"/> Sensitive – Security-Related-Periodic Review Required	<input type="checkbox"/> MD 3.4 Non-Public A.7 (Internal)
SUNSI Review Completed by: Jonathan Pfungsten		