

From: [Weidner, Tara](#)
To: [Jossian J Pagan Lisboa](#)
Subject: NRC request for additional information
Date: Friday, July 23, 2021 3:39:00 PM

***** PEASE CONFIRM RECEIPT OF THIS REQUEST AS
SOON AS POSSIBLE

SUBJECT: UNIVERSITY OF PUERTO RICO MEDICAL SCIENCES CAMPUS,
REQUEST FOR ADDITIONAL INFORMATION, MAIL CONTROL NO.
626933

Dear Jossian:

This is in reference to the letter dated May 10, 2021 requesting to amend NRC License No. 52-01946-07. In order to continue the review, I need the following additional information:

1. Please provide an e-mail address for Wanda Maldonado Dávila, Pharm D.
2. In your application it is unclear whether the area described as the Patient Suite and Injection Room is for the injection of PET patients only and if this is the room where PET patients rest prior to imaging. Please confirm what activities and patients will be using this room and if this is the only PET patient quiet room.
3. Confirm that you the lead cave designed to hold radioactive waste is adequate to shield PET radiopharmaceuticals and confirm that you possess a sharps container designed for PET as well.
4. Confirm that you will possess syringe shields and long handles forceps designed to handle PET radiopharmaceuticals.
5. Based on our independent shielding calculations, it appears that with the current thickness of concrete, you are not able to meet the public dose limit of 2 mrem in any one hour in the Anesthesia Surgical Supply Room (Third floor) and the roof above the Patient Suite and Injection Room. Please clearly describe how you intend to meet the regulatory requirement in 10 CFR 20.1301(a)(2).

I will continue my review upon receipt of this information. Please reply to my attention at: tara.weidner@nrc.gov.

In order to continue the review of your application, I request that you submit your response to this letter within 14 calendar days from the date of this e-mail.

Thanks in advance.

Tara L. Weidner, Senior Health Physicist
US NRC