

Seaman Nuclear Corporation

7315 South First Street Oak Creek, WI 53154 USA

Tel 414-762-5100 Fax 414-762-5106

info@seamannuclear.com

Leak Test Certificate

A leak test has been performed on meter, serial number **21004**, a model C-300 Non DT, containing the radionuclide Ra 226

Owned By: Milone & MacBroom Inc.
99 Realty Drive
Cheshire, CT 06410

Date Sample Collected: 12/23/20
Collected By: CB
Date Sample Analyzed: 2/11/21
Analyzed By: Andy Moore

Most regulatory agencies consider a source to be leaking if a leak test reveals the presence of more than 0.005 microcurie of removable contamination.

Analysis found contamination of less than 0.005 microcurie.

Analysis authorized by Wisconsin license 079-1257-01.

LEAK TEST DUE: 6/23/21



Scott C. Seaman
Radiation Safety Officer
Seaman Nuclear Corporation