

August 19, 2021

U.S. NRC Region III
2443 Warrenville Road
Suite 210
Lisle, Illinois 60532-4352

Licensee: Barnes-Jewish St. Peters Hospital
Address: 10 Hospital Drive; St. Peters, MO 63376
License #: 24-18968-01
Subject: **Response to Request for Additional Information**

Dear Sir or Madam,

Per email on 18 August 2021, the following information in offered:

As stated in the decommissioning report, the only radioisotopes used in the outpatient surgery center operating rooms were Pd-103 and I-125 permanent prostate implants. These are permanent brachytherapy seed implants in the patient's prostate. Any time a seed doesn't get implanted correctly, is leftover, or comes out of the patient's bladder, the seed is logged into the waste room and decayed prior to disposal. There were no temporary implants such as Cs-137 seeds. To answer the request for information, there was no leftover or excess material to be dispositioned.

Should you require additional information, please do not hesitate to contact me or Jeff Springer at jeffrey.springer@bjc.org.

Sincerely,



Constance Courtois
Radiation Safety Officer
St. Peters Hospital

From: [Cender, Laura](#)
To: [Pavon, Sandy](#)
Subject: File for ADAMS - Barnes-Jewish St. Peter's Hospital
Date: Thursday, August 26, 2021 12:03:47 PM
Attachments: [665 Response To Request for Additional Information.pdf](#)
[Response To Request for Additional Information.pdf](#)

Hello Sandy,

Please see the attached file to be entered into ADAMS for Barnes-Jewish St. Peter's Hospital. Please let me know if you have any questions.

Thank you,
Laura

Laura Cender
Pronouns: She/Her
U.S. Nuclear Regulatory Commission
Materials Licensing Branch
E-mail: Laura.Cender@nrc.gov
Phone: (630) 829-9712