



North East Technical Services, Inc.

75 Aileron Ct., Suite 4

Westminster, MD 21157

Ph: 410.751.5090 Fax: 410.751.5091

FMC & Associates, LLC

3927 Georgia Ave. NW, Ste C1

Washington, DC 20011

ATTN: Mely Mengstu 571-357-8098

*Shipping Address: 3927 Georgia Ave. NW, Ste C1
Washington, DC 20011*

LEAK TEST CERTIFICATE

MD Materials License # MD-13-020-01

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removeable contamination.

Gauge Model	3440	Gauge S/N	30578	Leak Test Date	4/16/2021
--------------------	-------------	------------------	--------------	-----------------------	------------------

Source

Reading in microCuries

750-5118

0.00008245

47-27680

0.00000

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by:

Douglas Sims

Date:

4/16/21

Monday, July 19, 2021