

**CERTIFICATE OF DISPOSITION
OF MATERIALS**

Estimated burden per response to comply with this mandatory collection request: 30 minutes. This estimate is used by NRC as part of the basis for its determination that the facility is released for unrestricted use. Send comments regarding burden estimate to the FOIA, Privacy, and Information Collections Branch (T-5 F53) U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001, or by e-mail to Information.Resources@nrc.gov, and to the Desk Officer, Office of Information and Regulatory Affairs, NECB-10232, (3150-0028), Office of Management and Budget, Washington, DC 20503. If a means used to impose an information collection does not display a currently valid OMB control number, the NRC may not contact or sponsor, and a person is not required to respond to, the information collection.

LICENSEE NAME AND ADDRESS

WOODIN & ASSOCIATES LLC
LEVELS BUSINESS PARK
111 PATRIOT DRIVE - SUITE D
MIDDLETOWN, DE 19709

LICENSE NUMBER

07-31039-01 030-36918

DOCKET NUMBER

LICENSE EXPIRATION DATE

APRIL 30, 2025

A. LICENSE STATUS (Check the appropriate box)

- This license has expired.
 This license has not yet expired; please terminate it.

B. DISPOSAL OF RADIOACTIVE MATERIAL

(Check the appropriate boxes and complete as necessary. If additional space is needed, provide attachments)

- The licensee, or any individual executing this certificate on behalf of the licensee, certifies that:
- 1. No radioactive materials have ever been procured or possessed by the licensee under this license.
 - 2. All activities authorized by this license have ceased, and all radioactive materials procured and/or possessed by the licensee under this license number cited above have been disposed of in the following manner.
 - a. Transfer of radioactive materials to the licensee listed below:
 - SEE ATTACHED LETTER
 - DATED 9-21-21.
 - Disposal of radioactive materials:
 - 1. Directly by the licensee:
 - 2. By licensed disposal site:
 - 3. By waste contractor:
 - c. All radioactive materials have been removed such that any remaining residual radioactivity is within the limits of 10 CFR Part 20, Subpart E, and is ALARA.

TETRA TECH
240 CONTINENTAL DRIVE-SURZOO
NEWARK, DE 19713
ATTN: RALPH BOEDEKER, P.E.
302/738-7551
NRC LICENSE No.: 25-29408-01

C. SURVEYS PERFORMED AND REPORTED

1. A radiation survey was conducted by the licensee. The survey confirms:
- a. the absence of licensed radioactive materials
 - b. that any remaining residual radioactivity is within the limits of 10 CFR 20, Subpart E, and is ALARA.
 - 2. A copy of the radiation survey results:
 - a. is attached; or
 - b. is not attached (Provide explanation); or
 - c. was forwarded to NRC on: _____ Date
 - 3. A radiation survey is not required as only sealed sources were ever possessed under this license, and
 - a. The results of the latest leak test are attached; and/or
 - b. No leaking sources have ever been identified.
- The person to be contacted regarding the information provided on this form:

NAME	TITLE	TELEPHONE (include Area Code)	E-MAIL ADDRESS
JOHN LESKO	SR. GEOTECHNICAL ENGR	302/602-3056	JLESKO@WOODIPASSOCIATES.COM
Mail all future correspondence regarding this license to:			
JOHN LESKO			

C. CERTIFYING OFFICIAL

I CERTIFY UNDER PENALTY OF PERJURY THAT THE FOREGOING IS TRUE AND CORRECT

PRINTED NAME AND TITLE	SIGNATURE	DATE
SENIOR GEOTECHNICAL ENGR/PM	<i>John Lesko</i>	9-22-21

WARNING: FALSE STATEMENTS IN THIS CERTIFICATE MAY BE SUBJECT TO CIVIL AND/OR CRIMINAL PENALTIES. NRC REGULATIONS REQUIRE THAT SUBMISSIONS TO THE NRC BE COMPLETE AND ACCURATE IN ALL MATERIAL RESPECT. 18 U.S.C. SECTION 1001 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.



September 21, 2021

Mr. John Lesko, PE
Woodin + Associates, LLC
Levels Business Park
111 Patriot Drive, Suite D
Middletown, DE 19709

Dear Mr. Lesko:

Subject: Receipt of Troxler Nuclear Density Gauge

Purpose of this letter is to document the transfer of ownership of Troxler Nuclear Density Gauge – Model 3430, Serial No. 36378 from Woodin & Associates, LLC to Tetra Tech, Inc. in the amount of \$10.00 – the check is attached. The density gauge was received from Wooding & Associates, LLC on September 21, 2021.

Sincerely,

A handwritten signature in cursive script that reads 'Ralph Boedecker'.

Ralph H. Boedecker, P.E.
Manager, Geotechnical Engineering and Construction Services



Troxler Electronic Laboratories, Inc.
 3008 Cornwallis Rd., P.O. Box 12057
 Research Triangle Park, NC 27709
 Tel: (877) 876-9637 Fax: (866) 391-2769
 License: NC 032-0182-1

JOHN LESKO
 WOODIN & ASSOCIATES
 111 PATRIOT DRIVE
 SUITE D
 MIDDLETOWN, DE 19709

Cust ID: 10646

LEAK TEST CERTIFICATE

DEVICE:

Model: 3430 Serial No: 36378

SEALED SOURCES:

Serial No.	Measure Date	Nuclide	GBq	mCi
78-1191	02/23/2005	Am-241:Be	1.48	40
77-3548	02/16/2005	Cs-137	0.296	8

LEAK TEST ANALYSIS:

Sample collected on: 04/07/2021
 Sample analyzed on: 04/14/2021 8:57:58 AM Position: 32
 Analyzed by: TH

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.23E+01	1.96E+01
Background measurement (cpm)	0	26
Sample measurement (cpm)	0	30
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	4.2E-01	1.3E+00

This certifies that the leak test results are:

Less than 185 Bq (0.005 uCi) Greater than 185 Bq (0.005 uCi)



Troxler Electronic Laboratories, Inc.
 3008 Cornwalls Rd., P.O. Box 12057
 Research Triangle Park, NC 27709
 Tel: (877) 876-9637 Fax: (866) 391-2759
 License: NC 032-0182-1

JOHN LESKO
 WOODIN & ASSOCIATES
 111 PATRIOT DRIVE
 SUITE D
 MIDDLETOWN, DE 19709

Cust ID: 10646

LEAK TEST CERTIFICATE

DEVICE:

Model: 3430 Serial No: 36378

SEALED SOURCES:

Serial No.	Measure Date	Nuclide	GBq	mCi
77-3648	02/16/2005	Cs-137	0.296	8
78-1191	02/23/2005	Am-241:Be	1.48	40

LEAK TEST ANALYSIS:

Sample collected on: 10/09/2020
 Sample analyzed on: 10/21/2020 10:20:08 A Position: 7
 Analyzed by: TH

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.23E+01	1.96E+01
Background measurement (cpm)	0	26
Sample measurement (cpm)	0	19
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	4.2E-01	1.3E+00

This certifies that the leak test results are:

Less than 185 Bq (0.005 uCi) Greater than 185 Bq (0.005 uCi)