

From: [Tran, Frank](#)
To: "supdike@mpcphysics.com"
Subject: Request for additional information for NRC License No. 13-01284-02
Date: Thursday, August 12, 2021 11:41:00 AM

Dear Ms. Updike:

We have reviewed the license amendment request dated July 21, 2021 for NRC License No. 13-01284-02 for Parkview Health. We will need the following.

1. The licensee requests to remove Item 6.H (Cs-137 sealed source for calibration) from the license. The licensee provided signed receipt, NRC Form 540; however, due to low resolution this document couldn't be read; please resubmit a readable copy of the form. Additionally, we couldn't find the last leak test for the sealed source Tech-OPS Model 77302 in the leak test record provided. Please submit a copy of the last leak test for Tech-OPS Model 77302.
2. The licensee stated that licensed material has not been used at Parkview North Medical Office Building, 11123 Parkview Plaza Dr., Ste. 201, Fort Wayne, Indiana. Please confirm that all licensed material and its residue, if any, has been removed from the location above.

To continue review of the application, we request that the licensee submits the response under a dated and signed cover letter by September 8, 2021. In the cover letter, please refer the license number, docket number and Mail Control No. 627442. If you have questions, please contact me at 630-829-9623 or reply to this email.

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this correspondence will be made available electronically for public inspection in the NRC Public Document Room or from the NRC's Agencywide Documents Access and Management System (ADAMS), accessible from the NRC Web site at <https://www.nrc.gov/reading-rm/adams.html>.

Sincerely,

Frank Tran

Health Physicist/License Reviewer
NRC Region III/Division of Nuclear Materials Safety
Phone: 630-829-9623
Fax: 630-515-1078
Email: Frank.Tran@nrc.gov

