



80 Seymour Street Hartford, CT 06106
24 September 2021

To: Robin Elliott
Senior Health Physicist
U.S. Nuclear Regulatory Commission
2100 Renaissance Boulevard, Suite 100
King of Prussia, PA 19406

From: Bette W. Blankenship M.S., D.A.B.R.
Medical Physicist and Radiation Safety Officer
Hartford Hospital
Department of Radiation Safety

RAM License No. 06-0025304

Dear Ms. Elliott;

In response to your email dated September 21st 2021 requesting additional source manufacturer and source specifics to add Cs-131 to the hospital prostate implant program:

Sealed source regulatory Information
Isoray Cs-131 Model CS-1
510(K): K030162
Regulation 21 CFR 892.5730
Radionuclide brachytherapy source
Regulatory Class II
Product code: 90 KXX
Approval received: March 28, 2003

Sealed Source and Device Registry Number: WA-1220-S-101-S
Approved September 17, 2004

Possession request: 2 Ci total activity, 10 mCi per seed maximum

Please contact me with any questions.

Thank you,

A handwritten signature in black ink that reads "Bette W. Blankenship". The signature is written in a cursive style with a large, prominent initial "B".

Bette W. Blankenship M.S., DABR, FAAPM
Medical Physicist and Radiation Safety Officer
Hartford Hospital
80 Seymour Street
Hartford, CT 06106
bette.blankenship@hhchealth.org