

**From:** [Forster, Sara](#)  
**To:** [Sheila.Shaffer@beaumont.org](mailto:Sheila.Shaffer@beaumont.org)  
**Subject:** Additional Information - Clarification Needed - re Beaumont Health Systems Amendment, Lic. No. 21-01333-01, CN628271  
**Date:** Thursday, September 02, 2021 4:19:00 PM

---

Good afternoon, Ms. Shaffer:

Our office has reviewed your August 18, 2021 request to amend the referenced license to remove the PET/CT radioactive materials use authorization for two locations in St. Clair Shores, Michigan. Based on our review, we understand that radioactive materials use at the St. Clair Shores Family Medicine Center, located at 21400 E Eleven Mile Rd., was limited to PET/CT radioactive materials use under Title 10 of the *Code of Federal Regulations* (CFR) Section 35.200. Could you please confirm that you wish to specifically remove the 21400 E Eleven Mile Rd. location of use (currently Condition Subitem No. 10.I. on the license) from the license, and that all radioactive materials authorized under the referenced license have been removed from that facility? If final closeout surveys are not available, due to continued use of the facilities at that location by another NRC licensee, could you please provide the most recent survey for that location, showing that the location is free of residual radiological contamination?

Please submit your response under a signed and dated letter. You may submit your response as a pdf file attached to an email message. If you have any questions regarding this message, please do not hesitate to contact me.

Thank you,

**Sara A. Forster, Health Physicist Licensing Reviewer**

U.S. Nuclear Regulatory Commission - Region III

Division of Nuclear Materials Safety

2443 Warrenville Rd. - Ste. 210

Lisle, IL 60532-4352

[sara.forster@nrc.gov](mailto:sara.forster@nrc.gov)

Direct: (630) 829-9892

Facsimile: (630) 515-1078

