



Laura Cender
U.S. Nuclear Regulatory Commission
2442 Warrenville Road
Suite 210
Lisle, IL 60532

October 18, 2021

Dear Ms. Cender,

I am writing to request an amendment to our NRC Materials License No. 13-23425-01. I wish to remove 535 Griswold Street, Detroit, MI 48826 as a storage location from our license due to upcoming office relocation. The gauges previously stored at this location are currently stored at other approved locations that are listed on our license. No gauges remain at this location. The only gauges previously stored at this location were Troxler Model 3440 serial numbers 20840 and 33289. These gauges have been leak tested twice yearly and have never been found to be leaking. I have enclosed copies of the most recent leak tests for these gauges for your reference.

Also, due to the upcoming office relocation noted above, I also need to change our licensee address from 535 Griswold Street, Detroit, MI 48826 to 625 Kenmoor Avenue SE, Suite 109, Grand Rapids, MI 49546.

Please let me know if you require any additional information. My email address is mzylstra@hntb.com and my phone number is 269-221-1678.

Sincerely,

A handwritten signature in black ink, appearing to read "Michael Zylstra". The signature is stylized and cursive.

Michael Zylstra, RSO
HNTB Corporation Michigan
625 Kenmoor Avenue SE, Suite 109
Grand Rapids, MI
49546



Troxler Electronic Laboratories, Inc.
 3008 Cornwallis Rd., P.O. Box 12057
 Research Triangle Park, NC 27709
 Tel: (877) 876-9537 Fax: (866) 391-2759
 License: NC 032-0182-1

MIKE ZYLSTRA
 HNTB
 625 KENMOOR AVE
 SUITE 109
 GRAND RAPIDS, MI 49546

Cust ID: 14444

LEAK TEST CERTIFICATE

DEVICE:

Model: 3440 Serial No: 20840

SEALED SOURCES:

Serial No.	Measure Date	Nuclide	GBq	mCi
47-16336	01/21/1992	Am-241:Be	1.48	40
75-2245	10/10/1991	Cs-137	0.296	8

LEAK TEST ANALYSIS:

Sample collected on: 06/28/2021
 Sample analyzed on: 07/12/2021 4:27:46 PM Position: 17
 Analyzed by: TH

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.23E+01	1.96E+01
Background measurement (cpm)	0	26
Sample measurement (cpm)	0	25
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	4.2E-01	1.3E+00

This certifies that the leak test results are:
 Less than 185 Bq (0.005 uCi) Greater than 185 Bq (0.005 uCi)



Troxler Electronic Laboratories, Inc.
 3008 Cornwallis Rd., P.O. Box 12057
 Research Triangle Park, NC 27709
 Tel. (877) 876-9537 Fax. (866) 391-2759
 License: NC 032-0182-1

THOMAS HUNDT
 HNTB
 38700 VAN DYKE AVE
 STE 100
 STERLING HEIGHTS, MI 48312

Cust ID: 14741

LEAK TEST CERTIFICATE

DEVICE:

Model: 3440 **Serial No:** 33289

SEALED SOURCES:

Serial No.	Measure Date	Nuclide	GBq	mCi
47-3345	12/17/1979	Am-241:Be	1.48	40
750-8671	09/28/2001	Cs-137	0.296	8

LEAK TEST ANALYSIS:

Sample collected on: 09/24/2021
Sample analyzed on: 09/29/2021 7:07:59 AM **Position:** 14
Analyzed by: TH

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.26E+01	1.99E+01
Background measurement (cpm)	0	26
Sample measurement (cpm)	0	13
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	4.1E-01	1.3E+00

This certifies that the leak test results are:

Less than 185 Bq (0.005 uCi) **Greater than 185 Bq (0.005 uCi)**

From: [Cender, Laura](#)
To: [Tomczak, Tammy](#)
Cc: [Pavon, Sandy](#); [Pavon, Martha](#)
Subject: FW: License Amendment Request HNTB License No 13-23425-01
Date: Monday, October 18, 2021 10:39:17 AM
Attachments: [Griswold Ammendment Letter HNTB License 13-23425-01.pdf](#)

Hello Tammy,

Please process the attached file in as a new license amendment request for HNTB Corporation.

Licensee: HNTB Corporation
License No. 13-23425-01
Docket No. 030-31268

Please let me know if you have any questions.

Thanks,
Laura

From: Michael Zylstra <mzylstra@HNTB.com>
Sent: Monday, October 18, 2021 7:05 AM
To: Cender, Laura <Laura.Cender@nrc.gov>
Subject: [External_Sender] License Amendment Request HNTB License No 13-23425-01

Laura,

Please see attached amendment request to remove 535 Griswold storage location from our license. We will be relocating out of this office in the new year. This also means we would need to change the license address. The information is included in the attached request letter. Please let me know if you require any additional information.

Thanks,

Mike

Michael Zylstra
Field Representative
Construction Services
Cell (269) 221-1678

HNTB CORPORATION
625 Kenmoor Avenue SE, Suite 109, Grand Rapids, MI 49546 | www.hntb.com

This e-mail and any files transmitted with it are confidential and are intended solely for the use of the individual or entity to whom they are addressed. If you are NOT the intended recipient and receive this

communication, please delete this message and any attachments. Thank you.