



CHRISTOPHER A. BOWENS, M.D., F.A.C.C.
APRN, FNP-BC

NOREEN MURPHY BOWENS, MSN,

1400 Peoples Plaza
Suite 111
Newark, DE 19702
302-834-3700

To Whom it May Concern,

Please see below the responses to the email sent from the Health Physicist of the US NRC requesting additional information in regards to our request to amend our license.

1. Please Identify whether any radioactive materials will be stored in the area labeled as "Injection Site", including waste.
 - a. No Radioactive materials will be stored in the area marked as "Injection Site"
2. Please identify any security measures that will be implemented to ensure no unauthorized access to radioactive materials will occur in the area labeled injection site.
 - a. All Radioactive materials are under direct supervision of the nuclear medicine technologist.
3. Please identify the surveys that will be performed in the area labeled as "Injection Site," including the frequency and acceptable limits.
 - a. Weekly wipes and daily surveys of the "Injection Site" including trash bins. Weekly wipes limits trigger level for unrestricted > 2200dpm/100 cm². Daily survey limits unrestricted trigger limits 0.5 mR/hr.

Sincerely,

A handwritten signature in cursive script that reads "Christopher Bowens MD".

Christopher A. Bowens, MD

October 29, 2021