

GHD Detroit, 26850 Haggerty Road
Farmington Hills, Michigan 48331
USA
www.ghd.com



Our ref: 11193079-006

11 November 2021

Materials Licensing Branch
USNRC Region III
2443 Warrenville Road, Suite 210
Lisle, IL 60532

Facility Removal Request from License No. 21-32548-01

Dear Laura Cender:

Per your request via e-mail on Monday, November 8, 2021, GHD Services Inc. (GHD) is submitting this letter notifying you that GHD no longer occupies the facility at the following address:

14496 Sheldon Road, Suite 200
Plymouth, Michigan 48170

In 2017, GHD relocated from the above Plymouth, MI facility to our current facility at 26850 Haggerty Road, Farmington Hills, Michigan. We are requesting that our Plymouth, Michigan facility be removed from our license. All licensed material has been removed as referenced in the previously submitted inventory record.

Also per your request, attached are the most recent leak test records for our two moisture density gauges (Serial Nos. 30968 and 31097) that were stored at the Plymouth, Michigan facility.

GHD believes the information provided herein satisfies the requests made via email on November 8, 2021. If you have any questions, please do not hesitate to ask.

Regards,

A handwritten signature in blue ink, appearing to read "Pete Swanson", is written over a light blue horizontal line.

Pete Swanson, PE
Principal-in-Charge

248-893-3407
pete.swanson@ghd.com



Troxler Electronic Laboratories, Inc.

3008 Cornwallis Rd., P.O. Box 12057
Research Triangle Park, NC 27709
Tel: (877) 876-9537 Fax: (866) 391-2759
License: NC 032-0182-1

BRENDON JUNGE
GHD SERVICES, Inc.
26850 HAGGERTY RD.
FARMINGTON HILLS, MI 48331

Cust ID: 13692

LEAK TEST CERTIFICATE

DEVICE:

Model: 3440 **Serial No:** 30968

SEALED SOURCES:

Serial No.	Measure Date	Nuclide	GBq	mCi
750-5473	08/05/1999	Cs-137	0.296	8
47-6307	06/17/1999	Am-241:Be	1.48	40

LEAK TEST ANALYSIS:

Sample collected on: 01/26/2021
Sample analyzed on: 02/10/2021 3:08:22 PM **Position:** 21
Analyzed by: TH

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.23E+01	1.96E+01
Background measurement (cpm)	0	26
Sample measurement (cpm)	0	21
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	4.2E-01	1.3E+00

This certifies that the leak test results are:

- Less than 185 Bq (0.005 uCi)** **Greater than 185 Bq (0.005 uCi)**



Troxler Electronic Laboratories, Inc.

3008 Cornwallis Rd., P.O. Box 12057

Research Triangle Park, NC 27709

Tel: (877) 876-9537 Fax: (866) 391-2759

License: NC 032-0182-1

BRENDON JUNGE
GHD SERVICES, Inc.
26850 HAGGERTY RD.
FARMINGTON HILLS, MI 48331

Cust ID: 13692

LEAK TEST CERTIFICATE

DEVICE:

Model: 3440

Serial No: 31097

SEALED SOURCES:

Serial No.	Measure Date	Nuclide	GBq	mCi
47-10731	10/05/1987	Am-241:Be	1.48	40
750-5736	08/24/1999	Cs-137	0.296	8

LEAK TEST ANALYSIS:

Sample collected on: 11/18/2020

Sample analyzed on: 11/23/2020 7:18:20 AM

Position: 2

Analyzed by: TH

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.23E+01	1.96E+01
Background measurement (cpm)	0	26
Sample measurement (cpm)	0	30
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	4.2E-01	1.3E+00

This certifies that the leak test results are:

Less than 185 Bq (0.005 uCi)

Greater than 185 Bq (0.005 uCi)

Pavon, Martha

From: Cender, Laura
Sent: Friday, November 12, 2021 11:37 AM
To: Pavon, Sandy; Pavon, Martha
Subject: File for ADAMS - GHD Services, Inc.
Attachments: GHD Facility Removal from License letter 11-11-2021.pdf

Hello,

Please enter the attached file into ADAMS for GHD Services, Inc.

Licensee: GHD Services, Inc.
License No. 21-32548-01
Docket No. 030-36768
Control No. 628705

Please let me know if you have any questions.

Thanks,
Laura

Laura Cender
Pronouns: She/Her
U.S. Nuclear Regulatory Commission
Materials Licensing Branch
E-mail: Laura.Cender@nrc.gov
Phone: (630) 829-9712