

**From:** [Gallagher, Robert](#)  
**To:** [Peter Mas](#)  
**Subject:** RE: Re: voice mail message  
**Date:** Tuesday, November 30, 2021 6:26:00 AM  
**Attachments:** [image003.png](#)

---

Good Morning Peter,

Would you send me documentation that Digirad has received that materials?

Thanks,  
Bob

---

**From:** Peter Mas <[pmas@snet.net](mailto:pmas@snet.net)>  
**Sent:** Monday, November 29, 2021 11:58 AM  
**To:** Gallagher, Robert <[Robert.Gallagher@nrc.gov](mailto:Robert.Gallagher@nrc.gov)>  
**Subject:** [External\_Sender] Re: voice mail message

I am re-sending it, just in case.

Thanks again for your help.

On Monday, November 29, 2021, 10:39:30 AM EST, Gallagher, Robert <[robert.gallagher@nrc.gov](mailto:robert.gallagher@nrc.gov)> wrote:

Did you send the 314 to me? I am still going thru my emails...

As for my "2<sup>nd</sup>" retirement, for now all I can say is sometime in 2022, perhaps summer but still not sure.

---

**From:** Peter Mas <[pmas@snet.net](mailto:pmas@snet.net)>  
**Sent:** Monday, November 29, 2021 10:36 AM  
**To:** Gallagher, Robert <[Robert.Gallagher@nrc.gov](mailto:Robert.Gallagher@nrc.gov)>  
**Subject:** [External\_Sender] Re: voice mail message

Got it - I can rely on an AMP to train an MD that is knowledgeable with HDR via the Elekta product and will now transition to using the Varian HDR device. License amendment required.

Thanks.

I sent over (maybe some 7-10 days ago) the Form 314 for New England Cardiology Associates in Manchester.

Hope your Thanksgiving was a grand event ! Stay well.

By the way, when are you going into your 2nd retirement? Spring 2022 ?

Sent from my iPhone

On Nov 29, 2021, at 10:25 AM, Gallagher, Robert <[Robert.Gallagher@nrc.gov](mailto:Robert.Gallagher@nrc.gov)> wrote:

Hi Peter,

10 CFR 35.690(c) states: "Has received training in device operation, safety procedures, and clinical use for the type(s) of use for which authorization is sought. This training requirement may be satisfied by satisfactory completion of a training program provided by the vendor for new users or by receiving training supervised by an authorized user or authorized medical physicist, as appropriate, who is authorized for the type(s) of use for which the individual is seeking authorization." We often see a checklist type of form used by Varian for each proposed user.

Bob

---

**From:** Peter Mas <[pmas@snet.net](mailto:pmas@snet.net)>  
**Sent:** Tuesday, November 23, 2021 5:28 PM  
**To:** Gallagher, Robert <[Robert.Gallagher@nrc.gov](mailto:Robert.Gallagher@nrc.gov)>  
**Subject:** [External\_Sender] Re: voice mail message

Thanks Robert-

Situation: An incoming physician who was authorized for HDR procedures with Ir-192 and most of their training/experience was on an Elekta (Nucletron) unit.

Question: What would be needed for them to become authorized to perform HDR but with the Varian GammaMed HDR unit that we have at Stamford ?

On Monday, November 22, 2021, 1:36:13 PM EST, Gallagher, Robert <[robert.gallagher@nrc.gov](mailto:robert.gallagher@nrc.gov)> wrote:

Hi Peter,

The voice mail message you left me was hard to understand. Can you reply to this email with your question?

Bob

Robert L. Gallagher

Health Physicist

Medical & Licensing Assistance Branch

U.S. NRC, Region I

2100 Renaissance Blvd.

King of Prussia, PA 19406

(610) 337-5182 office

(610) 337-5269 fax

