

From: [Hann, Patrick-John](#)
To: jmcilvain@dtcinc.com
Subject: Renewal Application- Request for Additional Information
Date: Tuesday, November 30, 2021 9:07:00 AM

This refers to your application dated July 28, 2021. In order to continue our review of your request, the following additional information is needed:

1. In your application dated July 28, 2021, you submitted shielding calculations and surveys for several nuclear medicine facilities, PET facilities, and an HDR treatment vault. Please identify the shielding materials, with thickness, that was installed in these facilities. Alternatively, you may confirm that these facilities were constructed with shielding that was specified in your shielding calculations.
2. In your application dated July 28, 2021, you submitted HDR emergency procedures. Does your facility model the manufacturer's procedures directly? If so, please provide a copy of the manufacturer's procedures.
3. In your application dated July 28, 2021, you submitted HDR spot check procedures. Please provide detailed, step-by-step procedures with the corresponding acceptance criteria.
4. In your application dated July 28, 2021, you stated that you have developed and will implement and maintain written procedures for sealed-source leak testing that meet the requirements of 10 CFR 35.67 and leak test sample collection and analysis will be performed by an organization authorized by the NRC or an Agreement State to provide leak testing services to other licensees; or by using a leak test sample collection kit supplied by an organization licensed by the NRC or an Agreement State to provide leak test kits or sample analysis services to other licensees and according to the instructions provided in the leak test sample collection kit. Please identify the contractor(s), with license number(s), that you will use for leak testing. Alternatively, you may confirm that you will perform in-house leak testing.
5. Please confirm that you wish to remove Dr. Sheldon and Mr. Walters from the license.

Your reply must be an originally signed and dated letter by management. The letter may be scanned and submitted as a pdf document attached to an email.

Patrick-John E. Hann, MHP
Health Physicist
U.S. Nuclear Regulatory Commission