

**CERTIFICATE OF COMPLIANCE
FOR RADIOACTIVE MATERIAL PACKAGES**

1.	a. CERTIFICATE NUMBER	b. REVISION NUMBER	c. DOCKET NUMBER	d. PACKAGE IDENTIFICATION NUMBER	PAGE	PAGES
	9290	9	71-9290	USA/9290/B(U)-85	1 OF	3

2. PREAMBLE

- a. This certificate is issued to certify that the package (packaging and contents) described in Item 5 below meets the applicable safety standards set forth in Title 10, Code of Federal Regulations, Part 71, "Packaging and Transportation of Radioactive Material."
- b. This certificate does not relieve the consignor from compliance with any requirement of the regulations of the U.S. Department of Transportation or other applicable regulatory agencies, including the government of any country through or into which the package will be transported.

3. THIS CERTIFICATE IS ISSUED ON THE BASIS OF A SAFETY ANALYSIS REPORT OF THE PACKAGE DESIGN OR APPLICATION

- | | |
|--|---|
| a. ISSUED TO (<i>Name and Address</i>)
Best Theratronics
413 March Road
Ottawa, Ontario
Canada K2K 0E4 | b. TITLE AND IDENTIFICATION OF REPORT OR APPLICATION
Best Theratronics application dated October 10, 2017,
as supplemented. |
|--|---|

4. CONDITIONS

This certificate is conditional upon fulfilling the requirements of 10 CFR Part 71, as applicable, and the conditions specified below.

5.

(a) Packaging

- (1) Model No. F-430/GC-40 Transport Package
- (2) Description

The Model No. F-430/GC-40 Transport package is designed to transport MDS Nordion's Gammacell-40 (GC-40) irradiator containing cesium-137 sealed sources in special form. The F-430 overpack provides impact and thermal protection for the radioactive contents. Containment is provided by the special form sealed source and shielding is provided by the GC-40 irradiator body.

The F-430 is stainless steel cylindrical package with a 50-inch-diameter and a height of 50 inches that is placed on a removable mild steel skid. The maximum weight of the package is 7000 pounds. The maximum weight of the GC-40 contents is 3835 pounds.

The overpack consists of nested cylindrical shells. The shells are made from stainless steel and the volume between the shells is filled with rigid foam. This foam provides insulation during an accidental fire. Vent holes, plugged with material designed to melt in a fire, are provided between the shells to prevent pressure buildup and allow a pathway for escape of gases from foam during an accidental fire.

**CERTIFICATE OF COMPLIANCE
FOR RADIOACTIVE MATERIAL PACKAGES**

1.	a. CERTIFICATE NUMBER	b. REVISION NUMBER	c. DOCKET NUMBER	d. PACKAGE IDENTIFICATION NUMBER	PAGE	PAGES
	9290	9	71-9290	USA/9290/B(U)-85	2 OF	3

The package contents consist of a Cesium-137 sealed source contained within an MDS Nordion GC-40 irradiator (upper or lower heads). The GC-40 is a research irradiator with lead shielding and a lead filled source drawer.

5.(a)(2) (continued)

The approximate dimensions and weights of the package are as follows:

Package outside diameter	50 inches
Package height	50 inches
Cavity diameter	36 inches
Cavity height	35.25 inches
Removable skid	50 inches x 50 inches x 8 inches (height)
Overpack weight	2640 pounds
Contents weight	3835 pounds
Maximum package weight	7000 pounds

(3) Drawings

The packaging is constructed in accordance with the Best Theratronics drawings F643001-001, Rev. P, sheet 1 of 3, and F643001-001, Rev. H, sheet 2 of 3, and F643001-001, Rev. B, sheet 3 of 3.

(b) Contents

(1) Type and form of material

Cesium-137 as a sealed source which meet the requirements of special form radioactive material.

(2) Maximum quantity of material per package

2,000 Curies.

6. In addition to the requirements of Subpart G of 10 CFR Part 71:

(a) The package must be prepared for shipment and operated in accordance with the Operating Procedures in Chapter 7 of the application.

(b) Each packaging must be acceptance tested and maintained in accordance with the Acceptance Tests and Maintenance Program in Chapter 8 of the application.

7. The package authorized by this certificate is hereby approved for use under the general license provisions of 10 CFR 71.17.

8. Revision No. 8 of this certificate may be used until February 28, 2022.

**CERTIFICATE OF COMPLIANCE
FOR RADIOACTIVE MATERIAL PACKAGES**

1.	a. CERTIFICATE NUMBER	b. REVISION NUMBER	c. DOCKET NUMBER	d. PACKAGE IDENTIFICATION NUMBER	PAGE	PAGES
	9290	9	71-9290	USA/9290/B(U)-85	3 OF	3


9. Expiration date: February 28, 2027.

REFERENCES

Best Theratronics application dated October 10, 2017.

Supplement dated December 7, 2021.

FOR THE U.S. NUCLEAR REGULATORY COMMISSION



Signed by Diaz-Sanabria, Yaira
on 01/17/22

Yaira K. Diaz Sanabria, Chief
Storage and Transportation Licensing Branch
Division of Fuel Management
Office of Nuclear Material Safety
and Safeguards

Date: January 17, 2022



Issuance of CoC No. 9290, Revision 9 DATE January 17, 2022

DISTRIBUTION:

KJamerson, NMSS/MSST/MSEB

ADimitriadis, R-I/DNMS/DIRHB

BDesai, R-II/DRS/EB3

DHills, R-III/DNMS/MCID

GWarnick, R-IV/DNMS/RIB

ADAMS Accession No.: ML22012A205; ML22012A208

OFFICE	NMSS/DFM/STLB	NMSS/DFM/STLB	NMSS/DFM/STLB	NMSS/DFM/CTCFB
NAME	BWhite <i>BW</i>	SFigueroa <i>SF</i>	BWhite <i>BW</i>	YDiaz-Sanabria <i>YD</i>
DATE	Jan 13, 2022	Jan 14, 2022	Jan 14, 2022	Jan 17, 2022

OFFICIAL RECORD COPY