



Principals
David I. Wiegand, P.G.
Charles R. Mitchell, P.E.
David S. Schultz, P.E.
Al Nouri, Ph.D., P.E.
John D. Woodburn, P.E.
Gerald S. Ross, Jr.
Ira L. Helms, P.G., P.E.

January 10, 2022

Mr. VonEhr,

Reference: 19-30397-01

Subject: Change of Conditions Request

Per Item 10 of our NRC license, I am writing to advise have closed our facility located at:

36 Chafee Lane
Fraziers Bottom, West Virginia

All gauges from that facility have been removed and taken to our main storage facility for our license which is located at:

4845 International Boulevard, Suite 104
Frederick, Maryland 21703

This letter is to request our license be amended to remove the Fraziers Bottom, West Virginia facility.

Should you have any questions please contact the undersigned directly at 301-514-8856 or via email at mbalcom@specializedengineering.com

Respectfully,

Mark Balcom
Corp. Dir. Of Quality Assurance



#4

North East Technical Services, Inc.
75 Aileron Ct., Suite 4
Westminster, MD 21157
Ph: 410.751.5090 Fax: 410.751.5091

Specialized Engineering - Frederick
4845 International Blvd., Suite 104
Frederick, MD 21703
ATTN: Mark Balcom

*Shipping Address: 4845 International Blvd, Suite 104
Frederick, MD 21703*

LEAK TEST CERTIFICATE

MD Materials License # MD-13-020-01

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removeable contamination.

Gauge Model	MC-3	Gauge S/N	M370903905	Leak Test Date	7/15/2021
	Source		Reading in microCuries		
	3905		0.00006557		
	3905		0.00000		

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by: *Douglas Sine*

Date: JUL 21 2021



#33

North East Technical Services, Inc.
75 Aileron Ct., Suite 4
Westminster, MD 21157
Ph: 410.751.5090 Fax: 410.751.5091

Specialized Engineering - Frederick
4845 International Blvd., Suite 104
Frederick, MD 21703
ATTN: Mark Balcom

Shipping Address: 4845 International Blvd, Suite 104
Frederick, MD 21703

LEAK TEST CERTIFICATE

MD Materials License # MD-13-020-01

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removeable contamination.

Gauge Model	3430	Gauge S/N	31408	Leak Test Date	7/15/2021
Source					
	750-6141		0.00006557		
	47-14615		0.00000		

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by: *Douglas Sines*

Date: **JUL 21 2021**