

**From:** [Elliott, Robin](#)  
**To:** [Jessica Trachten](#)  
**Cc:** [Denise Ferraiolo](#); [Indukala Doddamane](#); [Hann, Patrick-John](#)  
**Subject:** Request for additional information  
**Date:** Thursday, January 20, 2022 11:12:00 AM

---

License No.: 06-30624-01  
Docket No: 030-35657  
Control No: 628270

Licensee Name: Molecular NeuroImaging, LLC

Thank you for your January 12, 2022, response to our request for additional information. In order to continue our review of your request, the following additional information is needed:

The following item is referenced in the quoted section of NUREG 1556 Vol 9 Rev 3 Consolidated Guidance About Materials Licenses, Program Specific Guidance About Medical Use Licenses, <https://www.nrc.gov/reading-rm/doc-ollections/nuregs/staff/sr1556/v9/index.html>:

1. Relative to your response to our inquiry regarding Section 8.10.2 discussing Occupational Dose, you replied, "We will maintain, for inspection by the NRC, documentation demonstrating that unmonitored individuals are not likely to receive a radiation dose in excess of the limits in [10 CFR 20.1502](#)." Please confirm that this commitment is in addition to the previous information provided in your August 17, 2021, submission labeled Item 10: Occupational Dose.

The following item is referenced in the quoted section of NUREG 1556 Vol 7 Rev 1 Consolidated Guidance About Materials Licenses, Program Specific Guidance About Academic, Research and Development (R&D), and Other Licenses of Limited Scope, <https://www.nrc.gov/reading-rm/doc-collections/nuregs/staff/sr1556/v7/index.html>

1. Section 8.5.1 discusses Unsealed or Sealed Byproduct Material. Please provide the following for sealed sources, indicate if all the sources are used under 10 CFR 35.67 or if 10 CFR 30.32(g) applies. If the later, provide the required information as referenced in section 8.5.1 of NUREG 1556 Vol. & Rev 1.

Relative to your response regarding the distribution of radiopharmaceuticals, please either attach a copy of the October 27, 2011, letter or we will continue to tie down that letter separately in Condition 18.

Your reply must be an originally signed and dated letter by management. The letter may be scanned and submitted as a pdf document attached to an email (preferred); or it may be transmitted by facsimile to (610) 337-5269; or it may be sent by regular mail. In order to assure a prompt review of your renewal request, please reply no later than February 4, 2022. If you require additional time for your response or have questions regarding our request, please contact me.

Regards,

*Robin L. Elliott*

(Pronouns: she/her/hers)

Senior Health Physicist

Medical & Licensing Assistance Branch

Division of Radiological Safety and Security

U.S. NRC, Region I

2100 Renaissance Boulevard, Suite 100

King of Prussia, PA 19406-2713

(610) 337-5076 voice

(610) 337-5269 fax

[Robin.Elliott@nrc.gov](mailto:Robin.Elliott@nrc.gov)