


Good morning Jason,

1. We have one portable gauge. No incidents have occurred at the licensed permanent storage sites.
2. Gauge is in storage. I would like to sell the gauge. Going to reach out to Troxler.
3. We have a 3216 Roof moisture gauge from Troxler serial number 511. Last leak test was on 12-3-21 and is attached.
4. We will not use the gauge.
5. Gauge is in storage per the requirements.
6. Keri Wallace will be the RSO. No documentation is available at this point.
7. Gauge is in a secure storage and we will conduct an inventory on an annual basis
8. Gauge is in storage and will be selling shortly.
9. We will leak test at least once every 10 years and within 6 months before transfer to an authorized recipient.

Thank you,


Keri Wallace

From: [Kelly, Jason](#)
To: [Pavon, Sandy](#)
Cc: [Pavon, Martha](#)
Subject: FW: Signed
Date: Tuesday, April 12, 2022 9:21:25 AM
Attachments: [Good morning Jason..pdf](#)

Attached is an incoming response to a request for additional information for Materials License No. 21-32557-01 (Wallace & Associates, LLC), Docket No. 030-36875 and Control No. 629804.

Jason M Kelly, MPH
Health Physicist
U.S. NRC Region III – DNMS MLB
Phone: (630) 829-9737
E-mail: Jason.Kelly@nrc.gov

-----Original Message-----

From: Keri Wallace <keri@chenowethroofing.com>
Sent: Tuesday, April 12, 2022 9:17 AM
To: Kelly, Jason <Jason.Kelly@nrc.gov>
Cc: 'Ashley Dokes' <ashley@chenowethroofing.com>
Subject: [External_Sender] Signed

Jason,

Please see attached.

Thank you,

Keri Wallace