



Region I Office  
 Division of Nuclear Materials Safety  
 2100 Renaissance Boulevard, Suite 100  
 King of Prussia, PA 19406-2713  
 (610)337-5000

## Telephone Conversation Record

Date: May 10, 2022

License No. 39-19399-02

Docket No.(no hyphens): 03017547

Mail Control/Report No. 629247

Licensee Name: US Department of Commerce, NOAA

Participant(s) Name/Title: Ron Gagnon, RSO

Work Telephone No. 252-666-7468

Business Cellphone No. 252-259-5376

NRC Representative Name/Title: Randolph Ragland, Region I

Subject: US Dept. of Commerce, License Renewal, Request for Additional Information

Discussion: I informed Mr. Gagnon that NRC is in receipt of NOAA's Application dated 11/1/2021, and supporting information received on May 9, 2021, including a NUREG -1556, Vol. 7, Rev. 1 Appendix B Checklist and Radiation Safety Manual, but need additional information for NOAA's NRC license renewal. I requested the following:

- 1) Management signature
- 2) Statement to keep Chromium-51 or remove it from the license
- 3) Model Nos. for gas chromatographs containing Ni-63 and plans for dispositioning of these devices
- 4) Confirm that no financial assurance is required
- 5) If only invitro testing will be performed, confirm no animal testing
- 6) Please submit diagrams of areas where radioactive materials will be used or stored. Include a description of areas above and below and adjacent to these area and what materials will be used or stored in these areas.
- 7) Please commit that you will use survey instruments in accordance with Appendix I, NUREG -1556, Vol. 7, Rev. 1
- 8) If you want the flexibility to issue dosimetry or perform an evaluation, check the appropriate boxes in the checklist
- 9) Commit to performing sealed source leak tests at frequencies specified in the Sealed Source and Device Registry.
- 10) Either provide waste management procedures or commit to use the Appendix P, NUREG -1556, Vol. 7, Rev. 1 model waste procedures.
- 11) Please update your Radiation Safety Manual to exclude out-of-date information.

Action Required: Mr. Gagnon committed to provide the requested information.

<b>SUNSI REVIEW</b>			
<b>Document Availability:</b>	<input checked="" type="checkbox"/> Public	or	<input type="checkbox"/> Non-Public
<b>Document Sensitivity:</b> <i>(select "1" value to the right)</i>	<input checked="" type="checkbox"/> Non-Sensitive	<input type="checkbox"/>	MD 3.4 Non-Public B.1 (Non-Sensitive)
	<input type="checkbox"/> Non-Sensitive Copyright	<input type="checkbox"/>	MD 3.4 Non-Public A.3 (Sensitive Security Related)
	<input type="checkbox"/> Sensitive – Proprietary	<input type="checkbox"/>	MD 3.4 Non-Public A.4 (Sensitive Proprietary)
	<input type="checkbox"/> Sensitive – Privacy Act (includes Personally Identifiable Information (PII))	<input type="checkbox"/>	MD 3.4 Non-Public A.5 (Sensitive-Privacy Act)
	<input type="checkbox"/> Sensitive – Internal, Periodic Review required (All Other Sensitive Internal Info.)	<input type="checkbox"/>	MD 3.4 Non-Public A.6 (Sensitive-Federal, State, etc)
	<input type="checkbox"/> Sensitive – Security-Related-Periodic Review Required	<input type="checkbox"/>	MD 3.4 Non-Public A.7 ( Internal)
SUNSI Review Completed by: R. Ragland, May 9, 2022			