



May 16,2022

Shawn W. Seeley  
Health Physicist  
US Nuclear Regulatory Commission, Region 1  
2100 Reinassance Boulevard Suite 100  
King of Prussia  
Pennsylvania 19406-2713

Seeley, Shawn [shawn.seely@nrc.gov](mailto:shawn.seely@nrc.gov)

With regard to your additional questions:

1. In view of the history of the 31.11 authorization, we wish to have it removed from the license.
2. As far as the name variation it was a name change only which has not involved a change of control or ownership at the center.
3. We clarify that the ancillary and “research” personnel in our training section include includes AUs and NMTs.
4. We confirm that our instruments will be calibrated by an entity that is licensed by the NRC or an Agreement State.
5. We will develop, implement, and maintain written procedures for licensed material accountability and control to ensure that license possession limits are not exceeded; licensed material in storage is secured from unauthorized access or removal; licensed material, not in storage is maintained under constant surveillance and control; and records of receipt, transfer, and disposal of licensed material are maintained.



Francisco J Arraiza Antonmattei  
President