

**5/17/2022**

MATERIALS LICENSING BRANCH  
UNITED STATES NUCLEAR REGULATORY COMMISSION, Region III  
2443 Warrenville Road, Suite 210  
Lisle, IL 60532-4352

Re: Additional information concerning Control # 630878, Helen Newberry Joy Hospital, license # #21-32004-01

Dear Mr. Kelly:

Please find the additional information you requested below. We also have corrected the Area Survey information due to a typo.

**8.10.12: AREA SURVEYS**

"We have developed and will implement and maintain written procedures for area surveys in accordance with 10 CFR 20.1101 that meet the requirements of 10 CFR 20.1501 and 10 CFR 35.70."

**8.11 ITEM 11: WASTE MANAGEMENT**

"We have developed and will implement and maintain written waste disposal procedures for licensed material, in accordance with 10 CFR Part 20.1101, that also meet the requirements of the applicable section of 10 CFR Part 20, Subpart K, and of 10 CFR 35.92."

We will not dispose waste by incineration or compaction.

**Dr. Smith's information is as follows**

Dr. Zachary Smith, M.D. office phone 906-485-2632, cell phone 906-360-2985, fax 906-485-2622, and email [smith@upperpeninsulaimaging.com](mailto:smith@upperpeninsulaimaging.com).

Thank you for your assistance with this issue. If you have any further questions or require additional information, please contact our physics consultant, Laura Luna at (734)-662-3197, email at [lluna@mpcphysics.com](mailto:lluna@mpcphysics.com), or fax (734)-662-9224.

Sincerely,



**Hunter Nostrant, CEO**  
Helen Newberry Joy Hospital

**Your Health. Our Passion.**

## Pavon, Martha

---

**From:** Kelly, Jason  
**Sent:** Thursday, May 19, 2022 12:12 PM  
**To:** Pavon, Martha  
**Cc:** Pavon, Sandy  
**Subject:** FW: Control #630878  
**Attachments:** NRC Materials License Renewal Additional Information 2022.docx; Form 665 (Response to RAI Letter - 5-17-2022).pdf

Martha,

Good afternoon. Attached is an incoming Response to a Request for Additional Information from Materials License No. 21-32004-01 (Helen Newberry Joy Hospital), Docket No. 030-34481 for Control No. 630878.

Jason M Kelly, MPH  
Health Physicist  
U.S. NRC Region III – DNMS MLB  
Phone: (630) 829-9737  
E-mail: [Jason.Kelly@nrc.gov](mailto:Jason.Kelly@nrc.gov)

---

**From:** Holly Mischel <[holly.mischel@hnjh.org](mailto:holly.mischel@hnjh.org)>  
**Sent:** Wednesday, May 18, 2022 7:46 AM  
**To:** Kelly, Jason <[Jason.Kelly@nrc.gov](mailto:Jason.Kelly@nrc.gov)>  
**Cc:** [lluna@mpcphysics.com](mailto:lluna@mpcphysics.com)  
**Subject:** [External\_Sender] Control #630878

Hi Mr. Kelly,

Here's the additional information that you requested. If there's anything else you need please let me know. Thanks Holly

## Holly Mischel

*Radiology Director*

Helen Newberry Joy Hospital & Healthcare Center

502 W. Harrie Street, Newberry, MI 49868

P: 906.293.9102 F: 906.293.9135

E: [holly.mischel@hnjh.org](mailto:holly.mischel@hnjh.org)

---

**From:** Kelly, Jason [<mailto:Jason.Kelly@nrc.gov>]  
**Sent:** Friday, May 13, 2022 1:51 PM  
**To:** Holly Mischel <[holly.mischel@hnjh.org](mailto:holly.mischel@hnjh.org)>  
**Cc:** [lluna@mpcphysics.com](mailto:lluna@mpcphysics.com)  
**Subject:** [EXTERNAL]U.S. NRC Materials License No. 21-32004-01

**CAUTION:** This email originated from outside HNJJH.  
**DO NOT** click links or open attachments unless you recognize the sender and know the content is safe.

Ms. Mischel:

I have reviewed the request dated April 29, 2022, signed by Hunter Nostrant, CEO, for the renewal of U.S. Nuclear Regulatory Commission Materials License No. 21-32004-01.

The application was well prepared and in alignment with NUREG-1556, Volume 9, Rev. 3. Though, it appears that there was an inadvertent omission in Item 11 (Waste Management), in which the following statement appears:

We have developed and will implement and maintain written waste disposal procedures for licensed material, in accordance with 10 CFR Part 20 .1101 , that also meet the requirements of the applicable section of *10 CFR Part 20, Subpart* and of 10 CFR 35.92.

According to Section 8.11 of NUREG-1556, Volume 9, Rev. 3, an acceptable statement addressing Item 11 is:

We have developed and will implement and maintain written waste disposal procedures for licensed material in accordance with 10 CFR 20.1101, that also meet the requirements of the applicable section of *10 CFR Part 20, Subpart K*, and of 10 CFR 35.92.

Please resubmit Item 11 of your application, including the reference to Subpart K, which appears to have been inadvertently omitted. Please submit the correction under the cover of a letter bearing the signature of a management representative.

In the letter, please also include contact information for the proposed RSO, including office phone number, mobile phone number, fax number and e-mail address as applicable.

For routing purposes, please include a reference to U.S. NRC Materials License No. 21-32004-01 and Control No. 630878 in the letter. You may send me a scanned copy of the signed response letter via e-mail. Please provide the response within 15 calendar days of this letter so that I can expedite completion of your request.

**Jason M Kelly, MPH**  
**Health Physicist**  
**Materials Licensing Branch**  
**U.S. Nuclear Regulatory Commission**  
**Region III**  
**2443 Warrenville Road, Suite 210**  
**Lisle, IL 60532-4352**  
**Office: (630) 829-9737**  
**Fax: (630) 515-1078**



<http://www.nrc.gov/>

Privileged and Confidential: The information contained in this e-mail message, including any attachments, is intended only for the personal and confidential use of the intended recipient(s) and may contain confidential and privileged information as well as information protected by the Privacy Act of 1974. If the reader of this message is not the intended recipient or an agent responsible for delivering it to the intended recipient, you are hereby notified that you have received this document in error and that any review, dissemination, distribution, or copying of this message is strictly prohibited. If you have received this communication in error, please immediately contact the sender by reply e-mail and delete all copies of the original message.