



# Doctors Diagnostic Center

*Quality Diagnostic Testing*

May 19, 2022

United States

Nuclear Regulatory Commission

Region 1

210 Renaissance BLVD.

King of Prussia, PA 19406-2713

RE: Doctors Diagnostic Center, Request for additional information

NRC License No. 21-26766-01

To whom it may concern,

I, Mark Rosenberg, authorize Van Williams to make legally binding commitments on behalf of the licensee organization. Please contact me with any questions.

Thank You,

Mark Rosenberg, DO

(586) 264-1800

MR/kw

**2405 E. 14 Mile Rd.  
Sterling Heights, MI 48310  
Lower Level**

**Phone: 586-264-6079  
Fax: 586-264-2039**



UNITED STATES  
NUCLEAR REGULATORY COMMISSION  
REGION I  
2100 RENAISSANCE BLVD.  
KING OF PRUSSIA, PA 19406-2713

May 10, 2022

Van Williams, Manager  
Doctors Diagnostic Center  
2405 East 14 Mile Rd.  
Sterling Heights, MI 48310

SUBJECT: DOCTORS DIAGNOSTIC CENTER, REQUEST FOR ADDITIONAL  
INFORMATION, MAIL CONTROL NO. 630511

Dear Mr. Williams:

This is in reference to your application dated March 17, 2022, requesting to renew NRC License No. 21-26766-01. In order to continue our review, we need the following additional information. Please note that the specific requests and suggested format for responses to these items may be found in Appendix C to NUREG-1556, Vol. 9, Rev. 3, "Consolidated Guidance About Materials Licenses: Program-Specific Guidance About Medical Use Licenses" found at <https://www.nrc.gov/docs/ML1925/ML19256C219.pdf>:

1. The NRC views a letter signed by a management representative as indication that management has reviewed the application and concurs in the statements and representations contained therein. As such, please provide the following:
  - a. Please confirm that Mr. Van Williams is authorized to make legally binding commitments on behalf of the licensee organization.

OR

- b. If Mr. Williams is not authorized to make legally binding commitments on behalf of the licensee organization, please have a member of your organization authorized to make such commitments provide either of the following:
  - i. Confirm that you endorse the requests and statements submitted by Mr. Williams on behalf of your organization in the renewal request dated March 17, 2022.

OR

- ii. State that you do not endorse the requests and statements submitted by Mr. Williams on behalf of your organization in the renewal request dated March 17, 2022, and request to withdraw the requests and statements.

AND

- c. The delegation of authority letter dated March 17, 2022, was from Mr. Mark Rosenberg, D.O., but was signed by Mr. Williams. Please provide the following:
  - i. Provide a description of Dr. Rosenberg's position within the licensee's organization, including Dr. Rosenberg's title.

AND

- ii. Confirm that Mr. Williams signature was acceptable in place of Dr. Rosenberg's signature and is viewed as binding within the licensee's organization.
2. Item 5, Radioactive Material and Use - Please confirm you are not requesting to possess any additional sources that are not authorized under 10 CFR 35.65, 35.100, or 35.200.
3. Item 7, Radiation Safety Officer (RSO) or Associate Radiation Safety Officer (ARSO) – You are not required to submit the duties and responsibilities of the RSO; however, the list of typical RSO duties and responsibilities have been updated in Appendix I, 'Radiation Safety Officer Duties, Responsibilities, and Delegation' in NUREG-1556, Vol. 9, Rev. 3, 'Consolidated Guidance About Materials Licenses: Program-Specific Guidance About Medical Use Licensees.' No response is required for this item with respect to your license renewal application though it may be reviewed in a future inspection.
4. Item 10, Leak Tests – Your application contained the following commitment:

“Leak test sample collection and analysis will be performed by Medical Physics Consultants, Inc. (NRC License No. 21-20153-01) or another organization authorized by the NRC or an Agreement State to provide leak testing services to other licensees.”

Additional flexibilities are granted in the standard commitment found in NUREG-1556, Volume 9, Revision 3, “Consolidated Guidance About Materials Licenses: Program-Specific Guidance About Medical Use Licensees”. If desired, you may confirm and update your commitment to the following:

“Leak test sample collection and analysis will be performed by an organization authorized by the NRC or an Agreement State to provide leak testing services to other licensees; **or by using a leak test sample collection kit supplied by an organization licensed by the NRC or an Agreement State to provide leak test kits or sample analysis services to other licensees and according to the instructions provided in the leak test sample collection kit.**”

We will continue our review upon receipt of this information. Please reply to my attention at [Jonathan.Pfingsten@nrc.gov](mailto:Jonathan.Pfingsten@nrc.gov), referencing Mail Control number 630511.

In order to continue prompt review of your application, we request that you submit your response to this letter within 30 calendar days from the date of this letter.

An electronic version of the NRC's regulations is available on the NRC Web Site at: [www.nrc.gov](http://www.nrc.gov). Additional information regarding medical uses of radioactive materials may be obtained on the NRC Web Site at: <http://www.nrc.gov/materials/miau/med-use-toolkit.html>. This site also provides the updated Training and Experience NRC Form 313A series of forms and guidance, as well as information on the revised regulations for naturally-occurring and accelerator-produced radioactive materials (NARM).

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this letter will be available electronically for public inspection in the NRC Public Document Room or from the NRC's document system (ADAMS). ADAMS is accessible from the NRC Web Site at: <http://www.nrc.gov/reading-rm/adams.html>. Please be aware that you may request that certain portions of your submittal to NRC be withheld from public disclosure as proprietary information. To do this, you must execute an affidavit as specified in 10 CFR 2.390. You must list all portions that you wish to be held proprietary, along with your reasoning as to why that is appropriate. While it is allowable, please refrain from submitting proprietary information in support of a license unless necessary. Keep in mind that all NRC licenses are considered to be in the public domain, and therefore may be viewed by any member of the public who requests to see them.

If you have any questions regarding this request for additional information, please contact me at (610)337-5170 or via electronic mail at [Jonathan.Pfingsten@nrc.gov](mailto:Jonathan.Pfingsten@nrc.gov).

Thank you for your cooperation.

Sincerely,

**Jonathan B.  
Pfingsten**

Digitally signed by Jonathan  
B. Pfingsten  
Date: 2022.05.10 13:38:02  
-04'00'

Jonathan Pfingsten, Senior Health Physicist  
Medical and Licensing Assistance Branch  
Division of Radiological Safety and Security  
Region I

License No. 21-26766-01  
Docket No. 03034275  
Mail Control No. 630511

cc: Cari Dzanbazoff, Medical Physics Consultant  
John Mellen, M.D., Radiation Safety Officer

DOCTORS DIAGNOSTIC CENTER, REQUEST FOR ADDITIONAL INFORMATION, MAIL CONTROL NO. 630511 DATED MAY 10, 2022

DOCUMENT NAME: G:\WBL Documents\WBL License RAI\L21-26766-01.630511.RAI.docx

SUNSI Review Complete: Jonathan Pflingsten

After declaring this document "An Official Agency Record" it **will** be released to the Public.

To receive a copy of this document, indicate in the box: "C" = Copy w/o attach/encl "E" = Copy w/ attach/encl "N" = No copy

OFFICE	RIII:DNMS	N	RI:DRSS	N				
NAME	MGryglak/MRG		JPflingsten/JBP					
DATE	5/10/22		5/10/2022					

OFFICIAL RECORD COPY