

**From:** [Seeley, Shawn](#)  
**To:** [kstoneberg@ohiomedphys.com](mailto:kstoneberg@ohiomedphys.com)  
**Subject:** Davis Medical Center Request for Additional Information MC630612  
**Date:** Thursday, May 5, 2022 12:56:00 PM

---

Kelly,

In support of Davis Medical Center's (DMC) renewal application dated March 18, 2022, please provide the following additional information:

1. Your application has requested 35.200 materials. This now includes PET isotopes. Please confirm if PET materials will be utilized at DMC. If they will, please provide updated facility diagrams with the appropriate shielding calculations included. Please refer to "Consolidated Guidance About Materials Licenses: Program-Specific Guidance About Medical Use Licenses, Final Report (NUREG-1556, Volume 9, Revision 3)" for additional information on what to submit.
2. Your application did not adequately address material receipt and accountability. Either provide the statement found in the consolidated guidance (NUREG-1556, vol 9, rev 3), "We will develop, implement, and maintain written procedures for licensed material accountability and control to ensure that: license possession limits are not exceeded; licensed material in storage is secured from unauthorized access or removal; licensed material not in storage is maintained under constant surveillance and control; and records of receipt (either from the licensee's own production operations or from another licensee), transfer, and disposal of licensed material, are maintained," or you may provide your own procedures on how you will conduct material receipt and accountability under your license.
3. Your application did not address the leak test procedures of sealed sources under your license. Please provide the appropriate statement from NUREG-1556, vol 9, rev 3 for your response.

We will continue our review upon receipt of this information. Please let me know if you have any questions.

Shawn

*Shawn W. Seeley*

Shawn W. Seeley, Health Physicist  
USNRC  
2100 Renaissance Boulevard, Suite 100  
King of Prussia, PA 19406-2713  
610-337-5102 (o); 610-337-5269 (f)  
[Shawn.seeley@nrc.gov](mailto:Shawn.seeley@nrc.gov)