

**From:** [Seeley, Shawn](#)  
**To:** [crmidmr@aol.com](mailto:crmidmr@aol.com)  
**Subject:** Hospital Metropolitan Renewal Request for Additional Information MC630708  
**Date:** Thursday, May 5, 2022 1:11:00 PM

---

David,

In support of Hospital Metropolitan's (HM) renewal application dated April 5, 2022, please provide the following additional information:

1. Your application has requested 35.300 materials on an "as needed: basis." Please note that we require a maximum amount listed for this authorization. Your current license has a 1.5 Curie limit. Please confirm if this is the limit you wish to request.
2. Your application did not adequately address material receipt and accountability. Either provide the statement found in the consolidated guidance "Consolidated Guidance About Materials Licenses: Program-Specific Guidance About Medical Use Licenses, Final Report (NUREG-1556, Volume 9, Revision 3)", "We will develop, implement, and maintain written procedures for licensed material accountability and control to ensure that: license possession limits are not exceeded; licensed material in storage is secured from unauthorized access or removal; licensed material not in storage is maintained under constant surveillance and control; and records of receipt (either from the licensee's own production operations or from another licensee), transfer, and disposal of licensed material, are maintained," or you may provide your own procedures on how you will conduct material receipt and accountability under your license.
3. Your application did not adequately address the training for individuals working in or frequenting restricted areas. Please provide a statement that "We have developed and will implement and maintain written procedures for a program for training required under 10 CFR 19.12 for each group of workers, including (i) topics covered, (ii) qualifications of the instructors, (iii) method of training, (iv) method for assessing the success of the training, (v) initial training, and (vi) annual refresher training."
4. Your application did not address the leak test procedures of sealed sources under your license. Please provide the appropriate statement from NUREG-1556, vol 9, rev 3, in preparing your response.

We will continue our review upon receipt of this information. Please let me know if you have any questions.

Shawn

*Shawn W. Seeley*

Shawn W. Seeley, Health Physicist

USNRC

2100 Renaissance Boulevard, Suite 100

King of Prussia, PA 19406-2713  
610-337-5102 (o); 610-337-5269 (f)  
[Shawn.seeley@nrc.gov](mailto:Shawn.seeley@nrc.gov)