

October 18, 2022

United States Nuclear Regulatory Commission  
Licensing Division  
Region 1  
475 Allendale Road, Suite 102  
King of Prussia, PA 19406-1415

Subject: Franciscan Alliance, Inc., D/B/A Franciscan Health Crown Point, Response to Request for Additional Information

Dear Sir or Madam:

This letter confirms that I endorse the requests and statements submitted by Frank T. Godlewski, M.S., Radiation Safety Officer on behalf of Franciscan Alliance Inc. d/b/a/ Franciscan Health Crown Point in his letter dated October, 3, 2022 in the response to request for additional information letter dated September 22, 2022 regarding NRC Radioactive Material License #13-15933-01.

Sincerely,



Daniel McCormick, M.D.  
President/CEO

October 3, 2022

U.S. Nuclear Regulatory Commission  
Licensing Division  
Region I  
475 Allendale Road, Suite 102  
King of Prussia, PA 19406-1415

RE: Response to the request for additional information letter dated September 22, 2022, regarding NRC Radioactive Material License # 13-15933-01, Franciscan Health Crown Point


Dear Johnathan B. Pfingsten:

Please find the following responses to the items in the request for more information letter you sent to us:

1. b. i. Please find attached, the signed letter from FHCP President and CEO confirming endorsement of previous letters from Frank Godlewski.
2. a. The intended usage of current kits does not require authorization under 10 CFR 31.11 based on your statements.  
  
b. No actions required at this time.
3. a. i. Confirmed. Only sealed sources were ever possessed in the Radiation Oncology Hot Lab.  
ii. Please find attached, all the leak tests for the disposed Radiation Oncology Hot Lab sources  
iii. Confirmed. No sources stored in the Radiation Oncology Hot Lab have been identified as leaking.  
  
b. With respect to the radiation oncology hot lab close out document, the packaged sources and disposal paperwork are referring to the sources included on the waste manifests dated November 10, 2021.  
  
c. i. Ir-192 sources were not possessed under this license  
Cs-137 sources were possessed under this license.  
ii. Final date of clinical Use for Cs-137 = February 2014.  
iii. Cs-137 sources were disposed of with all other sources on Nov 10, 2021 manifest  
iv. Confirmed. The Cs-137 sources were not leaking. Please note attached leak tests.
4. At this point, we would like to retain Frank Godlewski as RSO for this renewal. We are considering appointing Joseph Yageski as ARSO, or RSO, depending on certain other conditions. Frank Godlewski, is and has been a consultant RSO for approximately the past 10 years at FHCP. On average, over that time, he has been onsite approximately 1 day per quarter and has provided oversight remotely of approximately 4 hours per week.

Thank you for your assistance in this matter. If you have any questions, please contact me at (312) 567-2193.

Sincerely,  
Franciscan Alliance, Inc,

A handwritten signature in black ink, appearing to be 'F. Godlewski', with a long horizontal line extending to the right.

Frank Godlewski  
Radiation Safety Officer