

**From:** [Robin Elliott \(She/Her\)](#)  
**To:** [Shanley, Ruth](#)  
**Subject:** Request for additional information  
**Date:** Wednesday, November 30, 2022 5:27:00 PM

---

License No.: 06-08544-01

Docket No: 030-01274

Control No: 633554

Licensee Name: Danbury Hospital

This refers to your request to amend your license dated October 31, 2022. In order to continue our review of your request, the following additional information is needed:

- Please provide the calibration data for both the Ludlum 14C GM meter and the Captus 3000 Well Counter used for the close out surveys of the Western CT Breast Imaging Center.
- Confirm that all seeds previously stored at this location have been moved to the 24 Hospital Avenue, Danbury, Connecticut address.
- Indicate whether the 95 Locust Avenue, Danbury, Connecticut should continue to be listed as a location for delivery of these seeds. If it is to be removed, confirm that packages were never opened and no packages remain at this location. If packages were opened at this location, a close out survey needs to be performed and submitted.

Your reply must be an originally signed and dated letter. The letter may be scanned and submitted as a pdf document attached to an email; or it may be transmitted by facsimile to (610) 337-5269. If we do not receive a reply from you within 30 calendar days from the date of this e-mail, we will assume that you do not wish to pursue your amendment request.

Please respond by e-mail to acknowledge that you have received the e-mail request for additional information.

Regards,

*Robin L. Elliott*

(Pronouns: she/her/hers)

Senior Health Physicist

Medical & Licensing Assistance Branch

Division of Radiological Safety and Security

U.S. NRC, Region I

475 Allendale Road, Suite 102

King of Prussia, PA 19406-1415

(610) 337-5076 voice

(610) 337-5269 fax

[Robin.Elliott@nrc.gov](mailto:Robin.Elliott@nrc.gov)

Training is useful, but there is no substitute for experience!

(From John Angle, M.D., ACMUI April 5, 2022 meeting)