

From: [Jason Kelly](#)
To: [Rachel Stieferman](#)
Subject: U.S. NRC Materials License No. 24-32390-01 - Request for Additional Information
Date: Thursday, January 19, 2023 5:31:00 PM

Ms. Stieferman:

I have reviewed the letter received December 9, 2022, and the accompanying revised application dated December 9, 2022, signed by Darryl Winegar, QC Director, in reference to the request to renew your U.S. Nuclear Regulatory Commission (NRC) Materials License No. 24-32390-01. Upon review of the response, I identified the following items where additional or clarifying information is needed:

1. Your license previously authorized 11 Seaman Nuclear Corporation Model C-200 and C-300 portable gauging devices. Though, your revised application identifies that you are requesting continuing authorization for only one Seaman Nuclear Corporation Model C-200 and C-300 portable gauging device. You may request to maintain authorization to possess, use or acquire the previously authorized 11 Seaman Nuclear Corporation Model C-200 and C-300 portable gauging devices.

Though, if you would like to reduce the license authorization, please clarify which model, either the C-200 or the C-300, that remains in your inventory. Further, please provide the following information for the Seaman Nuclear Corporation Model C-200 and C-300 portable gauging devices that have been transferred and/or disposed and that you would like continuing licensing authorization to possess. In addition, please provide applicable records documenting the disposal of the Seaman Nuclear Corporation Model C-200 and C-300 portable gauging devices.

- Submit records of transfer and/or disposal of the Seaman Nuclear Corporation Model C-200 or C-300 portable gauging devices formerly in your inventory. If returned to the manufacturer, please identify the date that the portable gauging devices were transferred/disposed. If transferred to another entity licensed by the U.S. NRC or an Agreement State, please identify the receiving entity, providing their U.S. NRC Materials License Number or a copy of their Radioactive Materials License if transferred to an entity in an Agreement State.
- Submit a copy of the most recent Leak Test Report for the Seaman Nuclear Corporation Model C-200 or C-300, which were performed ahead of the transfer/disposal. The Leak Test Reports should have an acceptable result identifying that there is no evidence of leakage.

2. Your license application included a copy of Appendix F, "Information Needed to Support Applicant's Request to Perform Nonroutine Maintenance," from NUREG-1556, Volume 1, Rev. 2, "Consolidated Guidance About Materials Licenses: Program-Specific Guidance About Portable Gauge Licenses." It appears that Appendix F was inadvertently included with the revised application. This is because your application identifies in the accompanying Appendix B that the gauge manufacturer or other person licensed by the NRC or an Agreement State will perform nonroutine maintenance or repair operations that require detaching the source or source rod from the gauge.

Please confirm that the inclusion of Appendix F was inadvertent. Or provide the applicable information specified in Appendix F to support the performance of nonroutine maintenance in-house.

Please respond to this request in writing within 15 calendar days. In your signed response letter, please reference your U.S. NRC Materials License No. 24-32390-01, Docket No. 030-35988 and Control No. 632059 for routing purposes. You are welcome to submit the response to me directly via e-mail at Jason.Kelly@nrc.gov.

Jason M Kelly, MPH
Health Physicist
Materials Licensing Branch
U.S. Nuclear Regulatory Commission
Region III
2443 Warrenville Road, Suite 210
Lisle, IL 60532-4352
Office: (630) 829-9737
Fax: (630) 515-1078



<http://www.nrc.gov/>