



CONVERSATION RECORD

NAME OF PERSON(S)/TITLE CONTACTED OR IN CONTACT WITH YOU		DATE OF CONTACT	TYPE OF CONVERSATION	
Joshua A. Young, M.S.		01/06/2023	<input type="checkbox"/> E-MAIL <input checked="" type="checkbox"/> TELEPHONE <input type="checkbox"/> INCOMING <input checked="" type="checkbox"/> OUTGOING	
E-MAIL ADDRESS		TELEPHONE NUMBER		
young127@purdue.edu		765-494-2721		
ORGANIZATION		DOCKET NUMBER(S)		
Purdue University Radiological and Environmental Management		030-00696		
LICENSE NAME AND NUMBER(S)		MAIL CONTROL NUMBER(S)		
Purdue University Radiological and Environmental Management License No. 13-02812-04		CN 633326		
SUBJECT				
Pending NRC License Amendment Request - Additional Information Required				
SUMMARY AND ACTION REQUIRED (IF ANY)				
Please provide your signed and dated response to the following items by no later than Friday, January 13, 2023.				
<ol style="list-style-type: none"> 1. Please clarify the specific brachytherapy source model(s) requested. Isoray currently has two cesium-131 sealed sources registered with the NRC; Cs-1 and Cs-1HA. 2. Please provide a statement confirming that if loose brachytherapy seeds are used for treatment that immediately following implantation a survey to locate and account for all sources that have not been implanted will be performed. 3. Please indicate if License Condition 30 should be amended to allow changes to the programs and procedures described in the letter dated October 27, 2022 that would meet requirements described in License Condition 30.A. - 30.D. 				
NAME OF PERSON DOCUMENTING CONVERSATION				
Laura B. Cender				
SIGNATURE			DATE OF SIGNATURE	
			01/06/2023	