

January 23, 2023

Attn: Document Control Desk
Director, Office of Nuclear Material Safety and Safeguards,
U.S. Nuclear Regulatory Commission,
Washington, DC 20555-0001.

RE: 2022 annual report of devices transferred for use and exempt from licensing under 10 CFR § 30.22 of or equivalent regulations of an Agreement State.

This report is being submitted in accordance with 10 CFR 32.32 and summarizes devices distributed by Viken Detection under NRC License Number 20-35671-01E. These devices have been transferred for use and exempt from licensing under 10 CFR § 30.22 of or equivalent regulations of an Agreement State.

The following Exempt Units were shipped in 2022:

Number of Units	Model	Description	Isotope	Activity (millicuries) per Unit
178	Pb200e	Handheld XRF used for Lead in Paint Testing	Co-57	5

These handheld XRF devices are authorized and described by Sealed Source and Device Registration NR-1397-D-102-E. If you have any questions, please do not hesitate to contact me by phone at 978-337-4517 or by email at jblute@vikendetection.com.

Regards,



Jim Blute
Viken Detection Corporation
21 North Avenue
Burlington, MA 01803