

**From:** [Jason Kelly](#)  
**To:** [Tennyson, Benny](#)  
**Cc:** [Norrick, Matt](#)  
**Subject:** U.S. NRC Materials License No. 13-26261-01  
**Date:** Monday, February 6, 2023 3:11:00 PM

---

Mr. Tennyson:

I have reviewed the request dated December 6, 2022, to amend your U.S. NRC Materials License No. 13-26261-01. Upon review of the request, I identified the following areas in which additional or clarifying information is needed:

Provide a copy of the Leak Test Report with an acceptable result to support the removal of the Instruments Inc., Division of National Tank Co., Model Y-1665 fixed gauging device. In lieu of a Leak Test Report, you may perform a closeout survey in the area where the device was previously used and/or stored. The closeout survey results should be keyed to a diagram depicting the survey points, background radiation level, identification of the radiation survey instrument (including calibration date), date that the radiation survey was performed and the name of the individual performing the survey.

Please respond to this request in writing within 15 calendar days. In your response, please refer to your U.S. Materials License No. 13-26261-01, Docket No. 030-32002 and Control No. 633817. Thank you.

**Jason M Kelly, MPH**  
**Health Physicist**  
**Materials Licensing Branch**  
**U.S. Nuclear Regulatory Commission**  
**Region III**  
**2443 Warrenville Road, Suite 210**  
**Lisle, IL 60532-4352**  
**Office: (630) 829-9737**  
**Fax: (630) 515-1078**



<http://www.nrc.gov/>