

**From:** [Jason Kelly](#)  
**To:** [Tennyson, Benny](#)  
**Cc:** [Norrick, Matt](#); [Pankey, Jim](#)  
**Subject:** RE: RE: U.S. NRC Materials License No. 13-26261-01  
**Date:** Monday, February 27, 2023 6:18:00 PM

---

Mr. Tennyson:

I have reviewed the closeout survey report. Upon review, I was not able to locate the following information:

- manufacturer and model number of the compatible probe used with your Ludlum Measurements, Inc., Model 3 survey meter;
- background radiation level;
- date that the closeout survey was performed;
- name of the individual performing the closeout survey;

Resubmit the closeout survey with the above requested information. I understand that the survey was performed following the removal of the Ra-226 sealed source. Please confirm that this is correct.

Please respond to this request in writing within 10 calendar days. In your response, please refer to your U.S. Materials License No. 13-26261-01, Docket No. 030-32002 and Control No. 633817. Thank you

**Jason M Kelly, MPH**  
**Health Physicist**  
**Materials Licensing Branch**  
**U.S. Nuclear Regulatory Commission**  
**Region III**  
**2443 Warrenville Road, Suite 210**  
**Lisle, IL 60532-4352**  
**Office: (630) 829-9737**  
**Fax: (630) 515-1078**



<http://www.nrc.gov/>

---

**From:** Tennyson, Benny <Benny.Tennyson@CountryMark.com>  
**Sent:** Friday, February 17, 2023 12:24 PM  
**To:** Jason Kelly <Jason.Kelly@nrc.gov>  
**Cc:** Norrick, Matt <Matt.Norrick@CountryMark.com>; Pankey, Jim <Jim.Pankey@CountryMark.com>

**Subject:** [External\_Sender] RE: U.S. NRC Materials License No. 13-26261-01

Jason

This is in reference to US Materials License No. 13-26261-01 Docket No. 030-32002 and Control No. 633817

Attached is drawings of the “old Radium Source Installation.

I spoke with RM Wester about a leak test report—at the present time they cannot provide this document

I performed a closeout survey. The results of that survey are also attached

I hope this is what you are needing

Benny

---

**From:** Jason Kelly <[Jason.Kelly@nrc.gov](mailto:Jason.Kelly@nrc.gov)>  
**Sent:** Monday, February 6, 2023 3:12 PM  
**To:** Tennyson, Benny <[Benny.Tennyson@CountryMark.com](mailto:Benny.Tennyson@CountryMark.com)>  
**Cc:** Norrick, Matt <[Matt.Norrick@CountryMark.com](mailto:Matt.Norrick@CountryMark.com)>  
**Subject:** U.S. NRC Materials License No. 13-26261-01

Mr. Tennyson:

I have reviewed the request dated December 6, 2022, to amend your U.S. NRC Materials License No. 13-26261-01. Upon review of the request, I identified the following areas in which additional or clarifying information is needed:

Provide a copy of the Leak Test Report with an acceptable result to support the removal of the Instruments Inc., Division of National Tank Co., Model Y-1665 fixed gauging device. In lieu of a Leak Test Report, you may perform a closeout survey in the area where the device was previously used and/or stored. The closeout survey results should be keyed to a diagram depicting the survey points, background radiation level, identification of the radiation survey instrument (including calibration date), date that the radiation survey was performed and the name of the individual performing the survey.

Please respond to this request in writing within 15 calendar days. In your response, please refer to your U.S. Materials License No. 13-26261-01, Docket No. 030-32002 and Control No. 633817. Thank you.

**Jason M Kelly, MPH**  
**Health Physicist**  
**Materials Licensing Branch**  
**U.S. Nuclear Regulatory Commission**  
**Region III**

**2443 Warrenville Road, Suite 210**

**Lisle, IL 60532-4352**

**Office: (630) 829-9737**

**Fax: (630) 515-1078**



**<http://www.nrc.gov/>**