DATE OF SIGNATURE

4/27/2023

SIGNATURE



CONVERSATION RECORD		
NAME OF PERSON(S)/TITLE CONTACTED OR IN CONTACT WITH YOU	DATE OF CONTACT	TYPE OF CONVERSATION
Robert F. Turco, Ph.D.	04/27/2023	E-MAIL
E-MAIL ADDRESS	TELEPHONE NUMBER	TELEPHONE INCOMING
robert.turco@mercy.net	(314) 251-6657	OUTGOING
ORGANIZATION	DOCKET NUMBER(S)	1
Mercy Hospital - St. Louis	030-02283	
LICENSE NAME AND NUMBER(S)	MAIL CONTROL NUMBER	(s)
Mercy Hospital - St. Louis 24-00794-03	CN 632241	
Pending NRC License Amendment Request - Additional Information Required SUMMARY AND ACTION REQUIRED (IF ANY)		
This request for information is in reference to the NRC license renewal request dated February 21, 2023. Please provide your response to the following items by no later than Friday, May 12, 2023.		
1. Provide a historical review of the radionuclides previously used in this area of use, including any sealed sources. If sealed sources were used in this area please provide a copy of the most recent leak test records and confirm that the sources have been transferred to another authorized location on the license.		
 During a recent quality assurance audit of licensin was determined that a required licensing commitment included in the submitted application and was missellowing statement: "We have developed and will implement and main 	nent relating to instrui sed during the review	ment calibration procedures was not v process. Please provide the
procedures in accordance with the requirements in CFR 35.61."	in 10 CFR 20.1501 a	nd that meet the requirements of 10
Additionally, it was identified that the submitted ap location as 1203 Smizer Mill Rd., St. Louis, MO bu Fenton, MO. Could you please confirm the correc	it the issued license I	lists the address as being located in
NAME OF PERSON DOCUMENTING CONVERSATION		
Laura B. Cender		

NRC Form 699 (11-2017) Page 1 of 1