



June 13, 2023

Jason M. Kelly, MPH
Health Physicist
U. S. Nuclear Regulatory Commission
Materials Licensing Section
2443 Warrenville Road, Suite 210
Lisle, IL 60532-4352

RE: Additional information

Mr. Kelly:

Please allow this letter to serve as a response to your request for additional information concerning the Deaconess Hospital c/o Midwest Radiologic Imaging amendment request, license number 13-00142-02. Your request for additional information was made via email and dated June 2, 2023.

We confirm that a change process similar to 10 CFR 35.26 will meet the following conditions:

- a. The revision is in compliance with the regulations;
- b. The revision is based upon NRC's current guidance for TheraSphere or SIR-Spheres Y-90 microspheres 35.1000 use posted on the NRC's Medical Uses Licensee Toolkit Web site;
- c. The revision has been reviewed and approved by the licensee's RSO and licensee's management;
- d. The affected individuals are instructed on the revised program before the change is implemented;
- e. The licensee will retain a record of each change for five years; and
- f. The record will include a copy of the appropriate website guidance, the old procedure, the new procedure, the effective date of the change, and the signature of the licensee management that reviewed and approved the change.

We confirm that we will also comply with the medical event reporting and notification requirements described in 10 CFR 35.3045(b)-(g).

If you have any questions concerning this submission, please feel free to contact Patrick Byrne at 734-717-8731 or pbyrne@mpcphysics.com.

Sincerely,

A handwritten signature in black ink that reads 'Susan Brumley'.

Susan Brumley
Vice President, Ancillary Services