

### SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

**1. LICENSEE/LOCATION INSPECTED**

Sibley Memorial Hospital  
Sibley Memorial Hospital  
5255 Loughboro Road, N.W.  
Washington, DC

**2. NRC/REGIONAL OFFICE**

Region 1  
475 Allendale Rd  
Suite 102  
King of Prussia, PA 19406

**REPORT NUMBER(S) 2023001**

**3. DOCKET NUMBER(S)**

030-14754

**4. LICENSE NUMBER(S)**

08-07398-03

**5. DATE(S) OF INSPECTION**

04/03/2023 - 07/17/2023

**LICENSEE:**

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC ) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.

\_\_\_\_\_ Non-cited violation(s) were discussed involving the following requirement(s):

- 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with the NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11. (Violations and Corrective Actions)

1: 10 CFR 35.70 requires, in part, that a licensee shall survey with a radiation detection survey instrument at the end of each day of use.

Contrary to this, the licensee failed to survey at the end of the day on five occasions. Specifically, the licensee did not make the required surveys on the following dates: March 4, 2023; March 6, 2023; March 17, 2023; March 20, 2023; and April 2, 2023.

This is a SLIV violation (6.3.d.3)

2: 10 CFR 20.1201 requires, in part, that the licensee control the occupational dose to individual adults.

License Condition No. 14 of NRC License Number 08-07398-03 requires, in part, that the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the enclosures listed. Thus includes the application dated December 4, 2014, which requires personnel who work in the restricted areas to wear a whole body dosimeter to monitor one's occupational dose.

Contrary to this, a cardiology authorized user was observed not wearing their dosimeter during a cardiac stress test.

Corrective actions included: setting a daily reminder in the NM hot lab to ensure daily surveys were performed and recorded before leaving for the day, and refresher training for all radiation workers and authorized users on the licensee's requirement for wearing dosimetry when working in the restricted areas.

### SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

**1. LICENSEE/LOCATION INSPECTED**

Sibley Memorial Hospital  
Sibley Memorial Hospital  
5255 Loughboro Road, N.W.  
Washington, DC

**2. NRC/REGIONAL OFFICE**

Region 1  
475 Allendale Rd  
Suite 102  
King of Prussia, PA 19406

**REPORT NUMBER(S) 2023001**

**3. DOCKET NUMBER(S)**

030-14754

**4. LICENSE NUMBER(S)**

08-07398-03

**5. DATE(S) OF INSPECTION**

04/03/2023 - 07/17/2023

#### Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

TITLE	PRINTED NAME	SIGNATURE	DATE
LICENSEE'S REPRESENTATIVE	Momodou Dibba	<i>Momodou L. Dibba</i>	07/18/2023
NRC INSPECTOR	Shawn Seeley	Shawn W. Seeley <small>Digitally signed by Shawn W. Seeley Date: 2023.07.18 14:49:34 -04'00'</small>	
BRANCH CHIEF	Christopher Cahill		