



Materials Inspection Report

1. Licensee/Location Inspected: Aggarwall & Associates M.D.'s, P.C. 15101 Southfield Road Allen Park, MI 48101 Report Number(s) 2023001	2. NRC/Regional Office Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352	
3. Docket Number(s) 030-36037	4. License Number(s) 21-32405-01	5. Date(s) of Inspection 05/10/2023, exit 07/27/2023

LICENSEE:
 The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

1. Based on the inspection findings, no violations were identified.

2. Previous violation(s) closed.

3. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements, and were assessed at Severity Level IV, in accordance with the NRC Enforcement Policy.

A. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy were satisfied.
 (Non-cited violation(s) was/were discussed involving the following requirement(s))

B. The following violation(s) is/are being cited in accordance with NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
 (Violations and Corrective Actions)

Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

TITLE	PRINTED NAME	SIGNATURE AND DATE
LICENSEE'S REPRESENTATIVE		
NRC INSPECTOR	Bill Lin	Bill C. Lin <small>Digitally signed by Bill C. Lin Date: 2023.07.26 09:47:52 -05'00'</small>
BRANCH CHIEF	Rhex Edwards	 <small>Digitally signed by Rhex A. Edwards Date: 2023.07.28 09:09:48 -05'00'</small>



Materials Inspection Record

1. Licensee Name: Aggarwall & Associates M.D.'s, P.C.	2. Docket Number(s): 030-36037	3. License Number(s) 21-32405-01
--	-----------------------------------	-------------------------------------

4. Report Number(s): 2023001	5. Date(s) of Inspection: 05/10/2023 exit 07/27/2023
---------------------------------	---

6. Inspector(s): Bill Lin	7. Program Code(s): 02201	8. Priority: 5	9. Inspection Guidance Used: 87130
------------------------------	------------------------------	-------------------	---------------------------------------

10. Licensee Contact Name(s): Rahel Makonnen	11. Licensee E-mail Address: rmakonn1@hfhs.org	12. Licensee Telephone Number(s): 734-671-1510
---	---	---

13. Inspection Type: <input type="checkbox"/> Initial <input checked="" type="checkbox"/> Routine <input type="checkbox"/> Announced <input type="checkbox"/> Non-Routine <input checked="" type="checkbox"/> Unannounced	14. Locations Inspected: <input type="checkbox"/> Hybrid <input type="checkbox"/> Main Office <input checked="" type="checkbox"/> Field Office <input type="checkbox"/> Temporary Job Site <input type="checkbox"/> Remote	15. Next Inspection Date (MM/DD/YYYY): 05/10/2028 <input checked="" type="checkbox"/> Normal <input type="checkbox"/> Extended <input type="checkbox"/> Reduced <input type="checkbox"/> No change
---	--	--

16. Location(s) Inspected List:
15101 Southfield Rd.
Allen Park MI 48101

17. Scope and Observations:
This was an unannounced, routine inspection of a private cardiology practice authorized to use byproduct material for diagnostic medical purposes at a clinic in Allan Park, Michigan. At the time of the inspection, the licensee employed one full-time technologist who performed approximately six cardiac stress tests per week. The technologist performed these studies on Fridays, starting at 9 am. The licensee received unit doses from a local radiopharmacy, and retained the services of a consulting physicist who audited the radiation safety program at least quarterly.

PERFORMANCE OBSERVATIONS

This inspection consisted of interviews with licensee personnel, a review of selected records, a tour of the nuclear medicine department in Allen Park, and independent measurements. The inspector could not observe any unit doses for cardiac stress test. At the initial visit with the licensee, the technologist was on vacation and then subsequently had additional issues that prevented the NRC inspector and the technologist to fully discuss the licensee's program until July 26, 2023. The inspector discussed with staff the implementation of licensee procedures for dose calibrator quality control, receipt of packages containing radioactive material, and area surveys. The inspector also reviewed documentation of dose administrations, area surveys, personnel dosimetry reports, and quarterly consultant audits.

No violations of NRC regulatory requirements were identified as a result of this inspection

Signature and Date - Branch Chief

Digitally signed by Rhex A. Edwards
Date: 2023.07.28 09:09:23 -05'00'