

From: [Shawn Seeley](mailto:Shawn_Seeley)
To: jhope@triadeng.com
Subject: Request for Additional Information MC635575
Date: Wednesday, August 16, 2023 11:32:00 AM

I am reviewing the renewal to your NRC license #47-31448-01 dated April 25, 2023. In order to complete our review, please submit the additional information requested below. Please let me know if you have any questions.

1. Please state the materials to be possessed and the proposed uses. Please include the radionuclide(s), the gauge manufacturer (and distributor) and model number(s), and the maximum number of gauges to be possessed. Also please confirm that the use will be in accordance with the applicable sealed source and device registry for each gauge.
2. Please submit a facility diagram for each storage location on your license. Include on the diagram the use of adjacent areas (including above and below), and information relevant to public dose and security as discussed in Sections 8.10.5, "Public Dose," and 8.10.6, "Operating, Emergency, and Security Procedures," respectively, in NUREG-1556, Vol. 1, Rev. 2, "Consolidated Guidance About Materials Licenses: Program-Specific Guidance About Portable Gauge Licenses"
3. Please confirm that you will develop, implement, and maintain procedures for ensuring accountability of licensed materials at all times.
4. Please confirm that the initial gauge users training meets the "Criteria" part of the section titled, "Training for Individuals Working in or Frequenting Restricted Areas" in NUREG-1556, Vol. 1, Rev. 2, "Consolidated Guidance About Materials Licenses: Program-Specific Guidance About Portable Gauge Licenses."
5. Please confirm that for "*Routine Cleaning and Lubrication*", you will implement and maintain procedures for routine maintenance of our gauges according to each manufacturer's written recommendations and instructions. Please confirm that for "*Nonroutine Maintenance*", you will have the gauge manufacturer or other person licensed by the NRC or an Agreement State perform nonroutine maintenance or repair operations that require detaching the source or source rod from the gauge.

We will continue our review upon receipt of this information. Please respond on company letterhead with a senior manager signature. You may respond directly via email to me.

Let me know if you have any questions.

Shawn
Shawn W. Seeley
Shawn W. Seeley
Health Physicist
USNRC

475 Allendale Road, Suite 102
King of Prussia, PA 19406
610-337-5102 (o); 610-337-5269 (f)
207-841-1157 (cell)
Shawn.seeley@nrc.gov