

From: [Shawn Seeley](#)
To: ghasems@mtigeo.com
Subject: Request for additional information MC635808
Date: Wednesday, August 16, 2023 10:58:00 AM

I am reviewing the renewal to your NRC license #45-35072-01 dated May 23, 2023. In order to complete our review, please submit the additional information requested below. Please let me know if you have any questions.

1. Please confirm that you wish to be authorized to use licensed material at temporary jobsites anywhere in the United States where the Nuclear Regulatory Commission maintains jurisdiction.
2. Please confirm that you will either possess and use, or have access to and use, a radiation survey meter that meets the criteria in the section titled, "Radiation Safety Program – Radiation Monitoring Instruments" in NUREG-1556, Vol 1, Rev.2, "Consolidated Guidance About Materials Licensees: Program Specific Guidance About Portable Gauge Licenses," in the event of an accident.
3. Please confirm that physical inventories will be conducted every 6 months or at intervals approved by the NRC to account for all sealed sources and devices received and possessed under the license, AND that you "Will develop, implement, and maintain procedures for ensuring accountability of licensed materials at all times."
4. Please confirm that you will maintain for inspection by the NRC, documentation demonstrating that unmonitored individuals are not likely to receive a radiation dose in excess of the limits in 10 CFR 20.1502(a), OR that you will provide and require the use of individual monitoring devices (dosimetry), and that all personnel dosimeters that require processing to determine radiation dose will be processed and evaluated by a NVLAP-approved processor.
5. Please confirm that you will implement and maintain the operating, emergency, and security procedures in Appendix G to NUREG-1556, Vol 1, Rev.2, "Consolidated Guidance About Materials Licensees: Program Specific Guidance About Portable Gauge Licenses," NUREG-1556, Vol 1, Rev.2, "Consolidated Guidance About Materials Licensees: Program Specific Guidance About Portable Gauge Licenses, ". Furthermore, please confirm that copies of the operating, security and emergency procedures will be provided to all gauge users and be available at each jobsite.
6. Please confirm that Leak tests will be performed at intervals approved by the NRC or an Agreement State and specified in the SSD registration certificate. Leak tests will be performed by an organization licensed by the NRC or an Agreement State to provide leak testing services to other licensees; or by using a leak test sample collection kit supplied by an organization licensed by the NRC or an Agreement State to provide leak test kits and/or sample analysis services to other licensees and according to the kit supplier's instructions. Records of leak test results will be

maintained.

7. Please confirm that for “*Routine Cleaning and Lubrication*”, you will implement and maintain procedures for routine maintenance of our gauges according to each manufacturer’s written recommendations and instructions. Please confirm that for “*Nonroutine Maintenance*”, you will have the gauge manufacturer or other person licensed by the NRC or an Agreement State perform nonroutine maintenance or repair operations that require detaching the source or source rod from the gauge.

We will continue our review upon receipt of this information. Please respond on company letterhead with a senior manager signature. You may respond directly via email to me.

Let me know if you have any questions.

Shawn

Shawn W. Seeley

Shawn W. Seeley

Health Physicist

USNRC

475 Allendale Road, Suite 102

King of Prussia, PA 19406

610-337-5102 (o); 610-337-5269 (f)

207-841-1157 (cell)

Shawn.seeley@nrc.gov