

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED

District Hospital Partners, L.P.
2150 Pennsylvania Avenue
Washington, DC 20037

2. NRC/REGIONAL OFFICE

Region 1
475 Allendale Rd
Suite 102
King of Prussia, PA 19406

REPORT NUMBER(S) 2023001

3. DOCKET NUMBER(S)

030-35424

4. LICENSE NUMBER(S)

08-30607-01

5. DATE(S) OF INSPECTION

04/04/2023 - 07/28/2023

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.

_____ Non-cited violation(s) were discussed involving the following requirement(s):

- 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with the NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11. (Violations and Corrective Actions)

1: 10 CFR 35.70 requires, in part, that a licensee shall survey with a radiation detection survey instrument at the end of each day of use.

Contrary to this, the licensee failed to survey at the end of the day on the date of the inspection. Specifically, the licensee made the required surveys at the start of the workday and not at the end of the workday as required.

This is a SLIV violation (6.3.d.3)

2: 10 CFR 20.1501 requires, in part, that the licensee make or cause to be made, surveys that are reasonable to evaluate the potential radiological hazards of radiation levels and residual radioactivity detected.

10 CFR 20.1906(b)(1) requires, in part, that the licensee shall monitor the external surfaces of packages for radioactive contamination.

Contrary to above, on April 5, 2023, the CNMT failed to survey the incoming package for radioactive contamination as required.

This is a SLIV violation (6.3.d.3)

The licensee took immediate corrective action and conducted training with all nuclear medicine technologists on the requirement to survey all areas at the end of the day, and to monitor all incoming packages for radioactive contamination. This was confirmed at the exit meeting held with licensee staff on July 28, 2023.

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED

District Hospital Partners, L.P.
2150 Pennsylvania Avenue
Washington, DC 20037

2. NRC/REGIONAL OFFICE

Region 1
475 Allendale Rd
Suite 102
King of Prussia, PA 19406

REPORT NUMBER(S) 2023001

3. DOCKET NUMBER(S)

030-35424

4. LICENSE NUMBER(S)


08-30607-01

5. DATE(S) OF INSPECTION

04/04/2023 - 07/28/2023

Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

| TITLE | PRINTED NAME | SIGNATURE | DATE |
|---------------------------|------------------|--|--|
| LICENSEE'S REPRESENTATIVE | Arnold Able |  | Digitally signed by Arnold Able Date: 2023.08.03 12:11:07 -04'00' |
| NRC INSPECTOR | Shawn Seeley | Shawn W. Seeley | Digitally signed by Shawn W. Seeley Date: 2023.08.03 12:17:41 -04'00' |
| BRANCH CHIEF | Anne DeFrancisco | | |