

From: [Magdalena Gryglak](#)
To: [Scott Surovi](#)
Subject: Request to change areas of use- NRC license no. 24-35692-01 for BJCMG Cardiology
Date: Tuesday, October 3, 2023 2:47:00 PM

Good afternoon,

I have reviewed your request dated September 19, 2023, to change the areas of use (addition of stress room no. 300.57).

I assume room 300.83 at 1520 Wentzville Parkway, Wentzville, MO 63385 is being released for unrestricted access.

Please provide a signed and dated letter with the following information (unless room 300.83 is not being released for unrestricted use at this time):

1. Please describe the unsealed radioactive material used in room 300.83 (list radioisotopes) and the date when the material was last used.
2. Please provide a list/inventory of sealed sources used in room 300.18 including serial numbers, quantities etc., if applicable.
3. Please provide the most recent leak test results for each of the sealed sources used in room 300.18, if applicable.
4. Please describe the location where the sealed sources were transferred. Please provide documentation acknowledging that all sealed sources (list of the sealed sources included) were received by an authorized recipient (another, waste disposal facility). Please provide a copy of the recipient's license (if not an NRC licensee).
5. A final status radiation survey and contamination survey conducted in room 300.83:
 - a. A survey map providing contamination (in cpm and dpm) and radiation readings (mR/hr) keyed to specific locations within the room (a diagram of the facility with cpm/dpm and mR/hr readings/points shown on the diagram with description of the reading/point such as floor, countertop, etc.);
 - b. Background contamination and radiation readings;
 - c. Guidance/action levels for contamination and radiation readings;
 - d. Instrumentation used to conduct the surveys including the manufacturer's names and model numbers;
 - e. Calibration dates for survey meters;
 - f. efficiency/correction factors for well counters;
 - g. Date and name/s of individuals performing the final status survey.
6. Confirmation that all radioactive waste has been decayed to background or disposed of to an authorized recipient, if applicable.
7. Confirmation that no radioactive material is present at the facility.

Please provide the information by October 31, 2023.

Thank you

Magdalena R. Gryglak
Health Physicist
Materials Licensing Branch
U.S. NRC Region III
630-829-9875