

To: Janice Nguyen, Senior Health Physicist  
Medical and Licensing Assistance Branch (MLAB)  
U.S. Nuclear Regulatory Commission, Region I  
475 Allendale Road, Suite 102  
King of Prussia, PA 19406-1415

19 January 2024

From: Windham Hospital  
112 Mansfield Avenue  
Willimantic, CT 06226

Subject: Radioactive Materials License Amendment Request (RAML 006-15203-01)

Dear Ms. Nguyen,

Kindly amend our Nuclear Regulatory Materials License (06-15203-01) as follows:

1. Please remove our old hot lab (Rm 1406 on floor plan attached) as an area of use. No radioactive materials, including radioactive source, have been stored in this room since 3 October 2023.

Attached, please find the survey results confirming that there is no fixed or removable contamination present in this area of use.

For any questions regarding this action please contact Erik Lyons, Asst. RSO at [erik.lyons@hhchealth.org](mailto:erik.lyons@hhchealth.org).

Sincerely,



Matthew Kaufman  
Vice President, Operations  
Hartford Healthcare East Region  
[Matthew.Kaufman@hhchealth.org](mailto:Matthew.Kaufman@hhchealth.org)

Windham Hot Lab Decommission Surveys. All measurements performed on 01/18/2024.

ROOM AND EQUIPMENT WIPES

ION CHAMBER- 451P-RYR

WELL COUNTER- Captus 4000e

SERIAL#: 8508

Serial#: 941303

Cal Date- 04/22/23

ROOM AND EQUIPMENT SURVEYS

**Wipe Background: 657 CPM    Exposure Background: 0.025mR/hr**


LOCATION	Area Surveys	Area Wipes
Lead Brick Area	0.011 mR/hr	-59 cpm
Hot Lab Sink	0.010 mR/hr	-72 cpm
Hot Lab Counters	0.011 mR/hr	11 cpm
Hot Lab Trash Area	0.011 mR/hr	-40 cpm
Hot Lab Sealed Source Storage Area	0.010 mR/hr	-40 cpm

Hot Lab Rm 1406



- Locations to Survey for Contamination**
- 1: Lead Brick cave
  - 2: Sink
  - 3: Counters
  - 4: Trash Storage area (under counter cabinets)
  - 5: Sealed source storage (under counter cabinets)

Hot lab (Rm 1406) identified as Radiopharmacy on drawing.





Designer and Manufacturer  
of  
Scientific and Industrial  
Instruments

www.ludlums.com

# CERTIFICATE OF CALIBRATION

**LUDLUM MEASUREMENTS, INC.**

501 Oak Street  
325-235-5494  
Sweetwater, TX 79556, U.S.A.



**CERT # 4084.01**

Customer WINDHAM HOSPITAL ORDER NO. 20444606/539277  
Mfg. Fluke Model 451P-RYR Serial No. 8508  
Mfg. \_\_\_\_\_ Model \_\_\_\_\_ Serial No. \_\_\_\_\_  
Cal. Date 22-Apr-23 Cal Due Date 22-Apr-24 Cal. Interval 1 Year Meterface Digital

Check mark  applies to applicable instr. and/or detector IAW mfg. spec. T. 73 °F RH 33 % Alt 709.0 mm Hg

- New Instrument  Instrument Received  Within Toler. +/-10%  10-20%  Out of Tol.  Requiring Repair  Other-See comments
- Mechanical ck.  Meter Zeroed  Background Subtract  Input Sens. Linearity  
 F/S Resp. ck.  Reset ck.  Window Operation  Geotropism  
 Audio ck.  Alarm Setting ck.  Batt. ck.  
 Calibrated in accordance with LMI SOP 14.8  Calibrated in accordance with LMI SOP 14.9

Instrument Volt Set \_\_\_\_\_ V Input Sens. \_\_\_\_\_ mV Det. Oper. \_\_\_\_\_ V at \_\_\_\_\_ mV Threshold Dial Ratio \_\_\_\_\_ = \_\_\_\_\_ mV  
 HV Readout (2 points) Ref./Inst. \_\_\_\_\_ / \_\_\_\_\_ V Ref./Inst. \_\_\_\_\_ / \_\_\_\_\_ V

**COMMENTS:**

Gamma Calibration: GM detectors positioned perpendicular to source except for M 44-9 in which the front of probe faces source.

Multimeter uncertainty within 1.3% of reading, Gamma uncertainty within 5.0% of reading, Neutron uncertainty within 7.0% of reading, Count rate uncertainty within 5.4% of reading

RANGE/MULTIPLIER	REFERENCE CAL. POINT	INSTRUMENT REC'D "AS FOUND READING"	INSTRUMENT METER READING
Digital			
Digital			
Digital			
Digital			
Digital			
Digital			
Digital			
Digital			
Digital			
Digital			

Range(s) Calibrated Electronically

REFERENCE CAL. POINT	INSTRUMENT RECEIVED	INSTRUMENT METER READING	Log Scale	REFERENCE CAL. POINT	INSTRUMENT RECEIVED	INSTRUMENT METER READING
Digital						
Readout	2 R/hr	1.85 R/hr				
	200 mR/hr	192 mR/hr				
	20 mR/hr	18.3				
	2 mR/hr	1.87 S				
	200 µR/hr	189 µR/hr				

Ludlum Measurements, Inc. certifies that the above instrument has been calibrated by standards traceable to the National Institute of Standards and Technology, or to the calibration facilities of other International Standards Organization members, or have been derived from accepted values of natural physical constants or have been derived by the ratio type of calibration techniques. All pass/fail determinations are based on the manufacturer's specifications without considering uncertainty factors. Measurement results represent expanded uncertainties expressed at approximately the 95% level of confidence, using a coverage factor of k=2. The calibration system conforms to the requirements of ANSI/NCSL Z540-1-1994 and ANSI N323A8-2013. ISO/IEC 17025:2017(E) State of Texas Calibration License No. LO-1963

Reference Instruments and/or Sources: Cs-137 S/N:  059  2171CP  2261CP  720  734  781  1131  1616  1696  1909  1916CP  2324/2521  
 5717CO  5719CO  60646  70897  73410  E552  G112  2168CP  S-394  S-1054  T10081  T10082 Neutron Am-241 Be  T-304 Ra-226  Y982  
 E551  5105  CSV280

Alpha S/N \_\_\_\_\_  Beta S/N \_\_\_\_\_  Other \_\_\_\_\_  
 m 500 S/N \_\_\_\_\_  Oscilloscope S/N \_\_\_\_\_  Multimeter S/N \_\_\_\_\_

Calibrator Donnie Miekos / Donnie Miekos Title Calibrator Date 22-Apr-23  
QC'd By [Signature] Title Final QC Date 24-Apr-23

This certificate shall not be reproduced except in full, without the written approval of Ludlum Measurements, Inc.  
FORM C22A 01/07/2020 Page 1 of 1

AC Inst.  Passed Dielectric (Hi-Pot) and Continuity Test  
Only  Failed: \_\_\_\_\_

Date: 1/4/2024

Nuclide: Cs137

Reference Activity: 0.5 uCi

Windham Hospital

Reference Date: 11/1/2023

Capintec 4000e SN#: 941303

SN: 2398-44-31

Minimum Detectable Activity Calculation (MDA)	
Nuclide: <b>Cs-137</b>	Counter Efficiency: <b>0.405</b>
Counts(N): <u>614</u>	Counting Time: <u>1</u> minutes
PF: <u>3.0</u>	CF: <u>4.65</u>
MDA: <u>291.91</u> dpm	$MDC_{(\text{gamma})} = \frac{3 + 4.65\sqrt{COUNT_{(\text{BACKGROUND})}}}{Efficiency * Counting Time}$
<u>0.000131</u> $\mu$ Ci	

MDA should be less than 0.005uCi for passing

Note: Counter Efficiency (CE) 30% = 0.300

Counts (Net) = Net Background Counts

Sample #		$[(N_i) - N_i \text{ Avg}]$	$[(N_i) - N_i \text{ Avg}] \times [(N_i) - N_i \text{ Avg}]$	Date: 1/4/2024
1	22952	256.50	65792.25	Instrument SN#: 941303  Nuclide: Cs137 Reference Activity: 0.5 uCi Reference Date: 11/1/2023 SN: 2398-44-31
2	22587	-108.50	11772.25	
3	22641	-54.50	2970.25	
4	22586	-109.50	11990.25	
5	22755	59.50	3540.25	
6	22726	30.50	930.25	
7	22754	58.50	3422.25	
8	22754	58.50	3422.25	
9	22674	-21.50	462.25	
10	22526	-169.50	28730.25	
$N_i \text{ Sum}$	226955			
$N_i \text{ Avg}$	22696			
		Sum	133033	
		CHI Square	5.86	

The probability is that it should fail 20% of the time: 10% high and 10% low. The Chi-Square result should be between 4.168 and 14.684.

1/4/2024

Windham Hospital

Capintec 4000e SN#: 941303

Nuclide: Cs137

Reference Activity: 0.5 uCi

Reference Date: 11/1/2023

SN: 2398-44-31

$$\text{Efficiency} = \frac{\text{Measured STD Activity (cpm)} - \text{bkg(cpm)}}{\text{Standard(uCi)} (2.22 \times 10^6 \text{ dpm/uCi})} * 100 =$$

	Nuclide	cpm	Activity (uCi)
Measured Standard Source	Cs-137	448200	
Measured Background		614	
Current Standard Source Activity:			0.4979

447586
1105338

% Efficiency
40.49