



Cardinal Health  
 Medical & Health Physics Services  
 5840 F Street  
 Omaha, NE 68117  
 402.734.8045 ext 6 tel  
 402.553.3033 fax

### CERTIFICATE OF LEAK TEST

**DATE:** 2/19/2024

**LICENSEE:** Digirad Imaging Solutions; Novi, MI

**COUNTING INSTRUMENT:** Cobra II Auto Gamma MCA

	SOURCE	TYPE	SERIAL NO	MODEL NO	ORIGINAL ACTIVITY	CALIBRATION DATE	CURRENT ACTIVITY	RESULTS
1	Co-57	Flood	BM07L100019057101	BM07L-10	10.00 mCi	3/28/2020	0.2653 mCi	<0.005 $\mu$ Ci

**COMMENT:** Sample collected by Stephanie Blanton.

**REMARKS:** The samples taken from each sealed source contained less than 0.005 microcuries of radioactive material. Each source meets the regulatory requirement for leak testing and scheduled testing.

**PERFORMED BY:**

Traci M. Jorgensen  
 For: Cardinal Health 414, LLC, Medical & Health Physics  
 State of Nebraska Registry Number: 50

**R.S.O.**

Stephanie Blanton

## Martha Pavon

---

**From:** Stephanie Blanton <Stephanie.Blanton@digirad.com>  
**Sent:** Wednesday, March 6, 2024 2:34 PM  
**To:** Bryan Parker  
**Subject:** [External\_Sender] RE: RE: RE: Please call asap  
**Attachments:** Digirad Novi MI.2024Feb19.pdf

Sorry for the delay on this – I've only just received the leak test back from Cardinal Health. Copy is attached. Let me know if you need anything else.

Thanks,  
Stephanie

---

**From:** Bryan Parker <Bryan.Parker@nrc.gov>  
**Sent:** Friday, February 16, 2024 2:13 PM  
**To:** Stephanie Blanton <Stephanie.Blanton@digirad.com>  
**Subject:** RE: RE: RE: Please call asap

**[EXTERNAL EMAIL]** This email was sent from outside of TTG Imaging Solutions. Please use caution opening attachments or clicking links.

Great, thanks for the update, Stephanie. I appreciate it.

Thanks.

*Bryan*

---

**From:** Stephanie Blanton <[Stephanie.Blanton@digirad.com](mailto:Stephanie.Blanton@digirad.com)>  
**Sent:** Friday, February 16, 2024 2:58 PM  
**To:** Bryan Parker <[Bryan.Parker@nrc.gov](mailto:Bryan.Parker@nrc.gov)>  
**Subject:** [External\_Sender] RE: RE: Please call asap

Looks like they missed it when doing leak tests – my guess is because the floods are stored in individual cases instead of with the vials and rod, that it was just forgotten. I'll get them to do a leak test and forward that on to you once received.

Thanks,

---

**From:** Bryan Parker <[Bryan.Parker@nrc.gov](mailto:Bryan.Parker@nrc.gov)>  
**Sent:** Friday, February 16, 2024 1:42 PM  
**To:** Stephanie Blanton <[Stephanie.Blanton@digirad.com](mailto:Stephanie.Blanton@digirad.com)>  
**Subject:** FW: RE: Please call asap

**[EXTERNAL EMAIL]** This email was sent from outside of TTG Imaging Solutions. Please use caution opening attachments or clicking links.

Hey Stephanie,

Left you a voicemail about this too – any word on this source? --- **see email below**

Just checking in so we can get this license terminated. Please let me know if you have any questions.

Thanks.

*Bryan*

**Bryan A. Parker**

Senior Health Physicist

USNRC Region III

[bryan.parker@nrc.gov](mailto:bryan.parker@nrc.gov)

678-828-7050

630-515-1078 (fax)



---

**From:** Bryan Parker

**Sent:** Friday, February 9, 2024 9:10 AM

**To:** Stephanie Blanton <[Stephanie.Blanton@digirad.com](mailto:Stephanie.Blanton@digirad.com)>

**Subject:** RE: RE: Please call asap

Hey Stephanie,

I am working on the termination for Dr. Gupta's license and had a quick question. On the info you sent me, a Co-57 source was transferred to you (nominal 10 mCi - s/n BM07L1019057101). But the leak tests provided only had the other 3 sources transferred (2 Cs-137 & 1 Ba-133). Do you have a leak test for that Co-57 source or can you give a status of it otherwise?

Please let me know if you have any questions.

Thanks.

*Bryan*

**Bryan A. Parker**

Senior Health Physicist

USNRC Region III

[bryan.parker@nrc.gov](mailto:bryan.parker@nrc.gov)

678-828-7050

630-515-1078 (fax)



---

**From:** Stephanie Blanton <[Stephanie.Blanton@digirad.com](mailto:Stephanie.Blanton@digirad.com)>  
**Sent:** Wednesday, January 10, 2024 3:53 PM  
**To:** Bryan Parker <[Bryan.Parker@nrc.gov](mailto:Bryan.Parker@nrc.gov)>  
**Subject:** [External\_Sender] RE: Please call asap

Good afternoon,

I've just gotten the signed 314 and termination letter back from Dr. Gupta. I've also included the close out survey that was performed back in August, along with the sealed source transfer paperwork and the most recent leak test for those sources. Please let me know if you need anything else.

Thanks,

**Stephanie Blanton**

Radiation Safety Officer | Digirad Imaging Solutions

P: (713)-376-1892

[stephanie.blanton@digirad.com](mailto:stephanie.blanton@digirad.com) | [www.digirad.com](http://www.digirad.com)



Delivering diagnostic imaging expertise. **As Needed. When Needed. Where Needed.**

---

**From:** Bryan Parker <[Bryan.Parker@nrc.gov](mailto:Bryan.Parker@nrc.gov)>  
**Sent:** Wednesday, December 27, 2023 9:43 AM  
**To:** Stephanie Blanton <[Stephanie.Blanton@digirad.com](mailto:Stephanie.Blanton@digirad.com)>  
**Subject:** Please call asap

**CAUTION:** This email originated from outside of the Digirad domain. Do not click links, open attachments, or reply unless you verify the sender's email and know the content is safe.

Ms. Blanton,

I left you a voicemail re: this issue about N. Gupta, MD and the status of their NRC license and your relationship with them under the Digirad license.

Dr. Gupta's license expires on 12/31 and we do not have a termination (or renewal request) for them. Your letter to the NRC dated 8/22/23 indicates that you would be assisting with the decommissioning and termination of that license.

Please call me as soon as possible so we can work toward resolving this matter.

Thanks.

*Bryan*

**Bryan A. Parker**

Senior Health Physicist

USNRC Region III

[bryan.parker@nrc.gov](mailto:bryan.parker@nrc.gov)

678-828-7050

630-515-1078 (fax)



**Email Confidentiality Notice:** The information contained in this transmission is confidential, proprietary or privileged and may be subject to protection under the law, including the Health Insurance Portability and Accountability Act (HIPAA) and Health Information Technology for Economic and Clinical Health (HITECH) Act. The message is intended for the sole use of the individual or entity to whom it is addressed. If you are not the intended recipient, you are notified that any use, distribution or copying of the message is strictly prohibited and may subject you to Federal and/or State law. If you received this transmission in error, please contact the sender immediately by replying to this email and delete the material from any computer.

## Martha Pavon

---

**From:** Bryan Parker  
**Sent:** Wednesday, March 6, 2024 3:30 PM  
**To:** Martha Pavon  
**Cc:** Sandy Pavon; Tammy Tomczak  
**Subject:** Please add AI  
**Attachments:** [External\_Sender] RE: RE: RE: Please call asap; 638532 AI 665.pdf

Hey Martha,

Please add the attached Additional information to CN638532 for N. Gupta, MD. A 665 is also attached.

Please let me know if you have any questions.

Thanks.

*Bryan*