

From: [Paul Jursinic \(PAJ Physics Consulting LLC\)](#)
To: [Ryan Craffey](#)
Cc: james.stopford@ascension.org
Subject: [External_Sender] Re: [EXTERNAL] FW: NRC Inspection Follow-up
Date: Thursday, March 7, 2024 7:19:23 AM

This message was sent securely using Zix®

Ryan,

The answer is yes to both. 1. The requirement for signatures of the written directive before the radioactive isotope is ordered is in our policy and procedures for Therasphere work. 2. The Authorized user and nuclear medicine staff have been trained in the policy and procedures for Therasphere work.

Paul Jursinic

On Thu, Feb 29, 2024 at 2:22 PM Ryan Craffey <Ryan.Craffey@nrc.gov> wrote:

Good afternoon Jim,

Please confirm that:

1. The process changes described in your written response have been incorporated into procedures.
2. All staff including nuclear medicine have been trained on the process changes.

Thank you!

Ryan Craffey

Senior Health Physicist

US Nuclear Regulatory Commission

Materials Inspection Branch, Region III

(630) 829-9655

--

Paul Jursinic, Ph.D.
Radiation Safety Officer
Ascension Health System (Borgess Hospital)