

From: Hiba Ahmed (She/Her)
Sent: Wednesday, March 13, 2024 3:46 PM
To: Miguel Vigo <nuclearvigo@gmail.com>
Cc: Robin Elliott (She/Her) <Robin.Elliott@nrc.gov>
Subject: Re: Bella Vista Hospital: Request for Additional Information (RAI)

Hello Dr. Vigo,

As a follow-up from our telephone conversation, we need the following information in order to process your renewal request.

1. In your submission, you have included the shielding calculations for above the uptake facility (quiet room) for your PET/CT laboratory. Please also provide shielding calculations and principle use for areas above the PET/CT imaging room and hot room as well.
2. In addition to the protective devices listed in your response, we wanted to confirm if a syringe shield carrier is utilized, specifically. Just to reiterate, you confirmed during our conversation that the doses are transported in a syringe shield and placed in a syringe carrier then transported on a cart.

Please respond to this request by Wednesday March 20th, 2024.
If you have any questions or concerns, please feel free to reach out.

Thank you,
Hiba Ahmed (she/her)
Health Physicist
Division of Radiological Safety and Security (DRSS)
Medical and Licensing Assistance Branch (MLAB)
U.S. Nuclear Regulatory Commission
Region I
Office (610) 337-5283
Hiba.ahmed@nrc.gov