

**COAL-MAC LLC**  
*1245 22 Mine Road*  
*Holden, WV 25625*  
*304/792-8400 (fax 304/792-8401)*

July 26, 2024

Mr. Randolph Ragland  
Senior Health Physicist  
**NRC Region I**  
475 Allendale Road, Suite 102  
King of Prussia, PA 19426

RE: Coal-Mac LLC (License No. 16-35383-01)  
Event 57121 (Fola Plant No. 2)


Dear Mr. Ragland,

During a leak test performed by Steve Ellis on 05-09-24 it was discovered that the shutter handle was not connected to the shutter column on the Ohmart Vega Model No. SH-F18-0 nuclear gauge that is located on the 2<sup>nd</sup> floor of Plant No. 2 near Bickmore, WV. This gauge houses 100mCi of Cs-137. It was determined that the shutter was indeed closed based on the radiation survey readings performed by Steve Ellis once he discovered that the shutter wasn't working properly. No individuals were exposed to high levels of radiation.

On 06-10-24 a technician from Vega Americas, Inc. performed a temporary fix on the unit to ensure that the shutter remains closed and confirmed that the shutter handle can be repaired by re-tapping the holes that connect the shutter handle. Vega permanently repaired the unit by re-tapping the holes on 07-23-24. This letter shall serve as the final update and Event 57121 should now be considered closed.

If you have any questions, please call me at (304) 792-8400.

Sincerely,



Doug Boyd,  
RSO  
Coal-Mac LLC