

**From:** [Andy Macias](#)  
**To:** [Zahid Sulaiman](#)  
**Cc:** [Larry \(JR\) Bradshaw](#)  
**Subject:** [External\_Sender] Incident Report # 57168  
**Date:** Monday, August 26, 2024 3:08:36 PM  
**Attachments:** [01 Missouri Uniform Crash Report.pdf](#)  
[02 S&A Performance Warning.pdf](#)  
[03 S&A Standard Operating Procedures Nuclear Gauge.pdf](#)  
[04 Radiation Safety Officer Certificate Larry Bradshaw.pdf](#)  
[05 NRC Incident #57168 Gauge Log-out Clarification.pdf](#)  
[Scan0144.pdf](#)

---

Mr. Zahid Sulaiman

Attached you will find the following information:

- 🔗 Highway patrol report
- 🔗 Employee performance warning
- 🔗 New standard operating procedures
- 🔗 Mr. Bradshaw's new RSO training certificate
- 🔗 Recent Qaltek service and maintenance report

In our initial incident report, Mr. Bradshaw stated that the gage was used on 6/14/24 , based on the log out book.

During your exit interview with Larry Bradshaw , it was noted that the leak test has a service date of 06-17-24. I spoke with Mr. Bradshaw regarding this issue, and the following is clarification of the gage use on Friday the 14th.

Mr. Bradshaw shared with me that leak test results were anticipated to be complete and reported on 06-14-24. Anticipating those results, Mr. Bradshaw logged out and secured gauge into work truck at 6:30a.m. on 06-14-24 for density testing on two projects as noted in the log out book. Mr. Bradshaw called at Qaltek at 11:05 am and he learned that getting the results on that day would not be possible. Mr. Bradshaw reported that no density testing was completed this day, 6/14/24. The gauge was not removed from the locked case, and the locked case not removed it from work truck. The gauge was logged back in to storage at 4:00pm. There was no exposure to public or workers on that day.

Leak test were provided by Qaltek on 06-17-24 , and first log out and actual density testing occurred 06-26-24.

I hope this clarifies the check out and use of the gage prior to leak test results.

Let us know if you have any questions.

Andy

**Andy Macias PE**  
**Branch Manager**

[Snyder & Associates, Inc.](#)

---

NOTICE: This E-mail (including all attachments) is covered by the Electronic Communications Privacy Act, 18 U.S.C. §§ 2510-2521, is confidential and may contain legally privileged information. If you are not the intended recipient, you are hereby put on notice that any retention, dissemination, distribution, or copying of this communication or any part thereof is strictly prohibited. Please call or reply to the sender immediately that you have received this message in error, then permanently delete it.