

Appendix

Consolidated Technical Analysis

The following table provides a list of 10 CFR Part 35 regulations and conditions the NRC has determined are applicable for use of *Eye90 microspheres*[®]. Licensees shall comply with all regulations which address use of *Eye90 microspheres*[®]. The table also provides specific conditions which the NRC has determined are necessary for the medical use of *Eye90 microspheres*[®]. In addition, the table lists where licensees and applicants can find additional guidance. Applicants may submit alternative list of regulations and specific conditions to be reviewed on a case-by-case basis by NRC staff.

Section	Description	Use Addressed in Regulation	Guidance	<i>Eye90 microspheres</i> [®] Guidance Section	Comment
35.1	Purpose and scope	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.2	Definitions	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.5	Maintenance of records	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.6	Provisions for the protection of human research subjects	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.7	FDA, other Federal, and State requirements	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		

Appendix

Consolidated Technical Analysis

Section	Description	Use Addressed in Regulation	Guidance	<i>Eye90 microspheres</i>[®] Guidance Section	Comment
35.8	Information collection requirements: OMB approval	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		

Appendix
Consolidated Technical Analysis

35.10	Implementation	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.11	License required	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.12	Application for license, amendment, or renewal	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.13	License amendments	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.14	Notifications	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.15	Exemptions regarding Type A specific licenses of broad scope	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.18	License issuance	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.19	Specific exemptions	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		

Appendix
Consolidated Technical Analysis

Subpart B – General Administrative Requirements					
35.24	Authority and responsibilities for the radiation protection program	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> Eye90 microspheres® <input type="checkbox"/> Other		
35.26	Radiation protection program changes	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> Eye90 microspheres® <input type="checkbox"/> Other		
35.27	Supervision	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> Eye90 microspheres® <input type="checkbox"/> Other		
35.40	Written directives (WDs)	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> Eye90 microspheres® <input type="checkbox"/> Other	6.2	35.40(b)(6) details written directive requirements for permanent implant brachytherapy. 35.40(b)(6)(ii) requires a record of the number of sources implanted, which is not possible due to the small size and large number of microspheres involved in Eye90 therapy. Therefore there are different items required by the written directive.
35.41	Procedures for administrations requiring a WD	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> Eye90 microspheres® <input type="checkbox"/> Other	6.2	The guidance document provides additional information specific to Eye90 Microspheres.
35.49	Suppliers for sealed sources or devices for medical use	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> Eye90 microspheres® <input type="checkbox"/> Other		
35.50	Training for Radiation Safety Officer (RSO) and Associate RSO	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> Eye90 microspheres®	5.2	The guidance document provides additional information specific to Eye90 Microspheres.

Appendix
Consolidated Technical Analysis

			<input type="checkbox"/> Other		
35.51	Training for an authorized medical physicist (AMP)	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> Eye90 microspheres® <input type="checkbox"/> Other		
35.55	Training for an authorized nuclear pharmacist (ANP)	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> Eye90 microspheres® <input type="checkbox"/> Other		
35.57	Training for experienced RSO, teletherapy or medical physicist, AMP, authorized user (AU), nuclear pharmacist, and ANP	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> Eye90 microspheres® <input type="checkbox"/> Other		
35.59	Recentness of training	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> Eye90 microspheres® <input type="checkbox"/> Other	5	The guidance document provides additional information specific to Eye90 microspheres®.
Subpart C – General Technical Requirements					
35.60	Possession, use, and calibration of instruments used to measure the activity of unsealed byproduct material	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> Eye90 microspheres® <input type="checkbox"/> Other	6.1	Although Eye90 microspheres® are not considered unsealed byproduct material, their pre-treatment activity is measured as per 35.60.
35.61	Calibration of survey instruments	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> Eye90 microspheres® <input type="checkbox"/> Other		
35.63	Determination of dosages of unsealed byproduct material for medical use	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> Eye90 microspheres® <input type="checkbox"/> Other	6.1	Although Eye90 microspheres® are not considered unsealed byproduct material, their pre-treatment activity is measured as per 35.60.

Appendix

Consolidated Technical Analysis

35.65	Authorization for calibration, transmission, and reference sources	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.67	Requirements for possession of sealed sources and brachytherapy sources	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other	6.5	Microspheres are not managed as individual discrete sources and Y-90 is short-lived. 35.406, 20.1906, 20.1904, 35.69, 20.1801, and 35.92 are sufficient.
35.69	Labeling of vials and syringes	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.70	Surveys of ambient radiation exposure rate	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other	6.8	The guidance document provides additional information specific to <i>Eye90 microspheres</i> [®] .
35.75	Release of individuals containing unsealed byproduct material or implants containing byproduct material	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.80	Provision of mobile medical service	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.92	Decay-in-storage	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other	7.3	The guidance document provides additional information specific to <i>Eye90 microspheres</i> [®] , including concerns specific to isotopic impurities inherent in the production of Y-90.
Subpart F – Manual Brachytherapy					

Appendix
Consolidated Technical Analysis

35.400	Use of sources for manual brachytherapy	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.404	Surveys after source implant and removal	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other	6.8	The guidance document provides additional information specific to Eye90 microspheres [®] . Eye90 microspheres [®] are permanent implants, therefore post-removal surveys are not applicable.
35.406	Brachytherapy sources accountability	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other	6.5	The guidance document provides additional information specific to Eye90 microspheres [®] .
35.410	Safety instruction	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.415	Safety precautions	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.432	Calibration measurements of brachytherapy sources	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other	6.1	Microspheres are too small to be calibrated in accordance with 35.432, therefore 35.63 and 35.60 are used for measuring each dosage.
35.433	Sr-90 sources for ophthalmic treatments	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.457	Therapy-related computer systems	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		

Appendix
Consolidated Technical Analysis

35.490	Training for use of manual brachytherapy sources	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other	5	The guidance document provides additional information specific to <i>Eye90 microspheres</i> [®] .
35.491	Training for ophthalmic use of Sr-90	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
Subpart L – Records					
35.2024	Records of authority and responsibilities for radiation protection programs	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.2026	Records of radiation protection program changes	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.2040	Records of WDs	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.2041	Records for procedure for administrations requiring a WD	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.2060	Records of calibrations of instruments used to measure the activity of unsealed byproduct materials	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1566 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.2061	Records of radiation survey instrument calibrations	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®]		

Appendix
Consolidated Technical Analysis

			<input type="checkbox"/> Other		
35.2063	Records of dosages of unsealed byproduct material for medical use	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> Eye90 microspheres® <input type="checkbox"/> Other		
35.2067	Records of leaks tests and inventory of sealed sources and brachytherapy sources	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> Eye90 microspheres® <input type="checkbox"/> Other	4.4, 6.5	Microspheres are not managed as individual discrete sources and Y-90 is short-lived. 35.406, 20.1906, 20.1904, 35.69, 20.1801, and 35.92 are sufficient.
35.2070	Records of surveys for ambient radiation exposure rate	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> Eye90 microspheres® <input type="checkbox"/> Other		
35.2075	Records of the release of individuals containing unsealed byproduct material or implants containing byproduct material	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> Eye90 microspheres® <input type="checkbox"/> Other		
35.2080	Records of mobile medical services	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> Eye90 microspheres® <input type="checkbox"/> Other		
35.2092	Records of decay-in-storage	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> Eye90 microspheres® <input type="checkbox"/> Other	7.3	The guidance document provides additional information specific to Eye90 microspheres®.
35.2204	Records of Mo-99, Sr-82, and Sr-85 concentrations	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> Eye90 microspheres® <input type="checkbox"/> Other		

Appendix

Consolidated Technical Analysis

35.2404	Records of surveys after source implant and removal	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other	6.8	The guidance document provides additional information specific to <i>Eye90 microspheres</i> [®] . <i>Eye90 microspheres</i> [®] are permanent implants, therefore post-removal surveys are not applicable.
35.2406	Records of brachytherapy source accountability	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other	6.5	<i>Microspheres</i> are not managed as individual discrete sources and Y-90 is short-lived. 35.406, 20.1906, 20.1904, 35.69, 20.1801, and 35.92 are sufficient.
35.2432	Records of calibration measurements of brachytherapy sources	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other	6.1	<i>Microspheres</i> are too small to be calibrated in accordance with 35.432, therefore 35.63 and 35.60 are used for measuring each dosage.
35.2433	Records of decay of Sr-90 sources for ophthalmic treatments	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
Subpart M – Reports					
35.3045	Report and notification of a medical event	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other	6.3	The guidance document provides additional information specific to <i>Eye90 microspheres</i> [®] .
35.3047	Report and notification of a dose to an embryo/fetus or a nursing child	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.3067	Report of a leaking source	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.3204	Report and notification for an eluate exceeding permissible Mo-99, Sr-	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®]		

Appendix
Consolidated Technical Analysis

	82, and Sr-85 concentrations		<input type="checkbox"/> Other		
Subpart N – Enforcement					
35.4001	Violations	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.4002	Criminal penalties	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
Additional Considerations					