

**From:** [Bradley Smith](#)  
**To:** [Tony Gonzalez](#)  
**Cc:** [Jason Webb](#)  
**Subject:** [External\_Sender] RE: RE: Radioactive Materials License Renewal - Request for Information  
**Date:** Monday, February 3, 2025 9:19:26 AM

---

Hello Tony,

Thank you for the reminder, I have been traveling a lot and struggling to keep up. In regard to our Liquid Scintillation Counter, we calibrate annually as to the following statement:

From NUREG 1556 Vol. 12 Appendix B for Instrument Calibration:

State that: "Instruments will be calibrated before first use, at least annually thereafter, and after any repair, by a vendor that the NRC or an Agreement State has licensed to perform instrument calibration."

**Thanks,**  
**Bradley R. Smith**  
*Director of Engineering/RSO*

**HIVIZ Shooting Systems**  
**620 S. Adams Street | Laramie, WY 82070**  
**Phone 307.223.2785 | Toll Free: 800.589.4315**  
[brad@hivizsights.com](mailto:brad@hivizsights.com) | [www.HiVizSights.com](http://www.HiVizSights.com)

---

**From:** Tony Gonzalez <Antonio.Gonzalez@nrc.gov>  
**Sent:** Monday, February 3, 2025 8:12 AM  
**To:** Bradley Smith <brad@hivizsights.com>  
**Subject:** FW: RE: Radioactive Materials License Renewal - Request for Information

Hi Brad,

I'm just resending this in case you missed it. Please let me know your response to the below.

Thank you,  
Tony

---

**From:** Tony Gonzalez  
**Sent:** Friday, January 24, 2025 9:37 AM  
**To:** Bradley Smith <[brad@hivizsights.com](mailto:brad@hivizsights.com)>  
**Subject:** RE: RE: Radioactive Materials License Renewal - Request for Information

Hello Brad,

With regard to your Liquid Scintillation Counter, what does HiViz commit to?

From NUREG 1556 Vol. 12 Appendix B for Instrument Calibration:

State that: “Instruments will be calibrated before first use, at least annually thereafter, and after any repair, by a vendor that the NRC or an Agreement State has licensed to perform instrument calibration.”

**OR**

State that: “We will implement the model radiation survey instrument calibration program published in Appendix H in NUREG–1556, Volume 12, Rev. 1, “Consolidated Guidance About Materials Licenses: Program-Specific Guidance About Possession Licenses for Manufacturing and Distribution.””

**OR**

Submit equivalent procedures for instrument calibrations.

Thank you,  
Tony

---

**From:** Bradley Smith <[brad@hivizsights.com](mailto:brad@hivizsights.com)>  
**Sent:** Friday, December 20, 2024 1:58 PM  
**To:** Tony Gonzalez <[Antonio.Gonzalez@nrc.gov](mailto:Antonio.Gonzalez@nrc.gov)>  
**Cc:** Phil Howe <[phil@hivizsights.com](mailto:phil@hivizsights.com)>; Jason Webb <[Jason@hivizsights.com](mailto:Jason@hivizsights.com)>; Timothy Brandon <[tbrandon@diligistics.com](mailto:tbrandon@diligistics.com)>  
**Subject:** [External\_Sender] RE: Radioactive Materials License Renewal - Request for Information

Hello Mr. Gonzalez,

Please find the attached response letter along with updated tritium floor plans for HIVIZ.  
Thanks for your work on this renewal.

**Thanks,**  
**Bradley R. Smith**

*Director of Engineering/RSO*

**HIVIZ Shooting Systems**

**620 S. Adams Street | Laramie, WY 82070**

**Phone 307.223.2785 | Toll Free: 800.589.4315**

**[brad@hivizsights.com](mailto:brad@hivizsights.com) | [www.HiVizSights.com](http://www.HiVizSights.com)**

---

**From:** Tony Gonzalez <[Antonio.Gonzalez@nrc.gov](mailto:Antonio.Gonzalez@nrc.gov)>

**Sent:** Thursday, December 12, 2024 11:22 AM

**To:** Bradley Smith <[brad@hivizsights.com](mailto:brad@hivizsights.com)>

**Subject:** Radioactive Materials License Renewal - Request for Information

Good morning Brad,

I am in receipt of your request for license renewal. However, to complete the review I need some additional information I have laid out in the attached letter. Please review the attached and provide a response within the requested time period. If you have any questions please feel free to call me at 817-200-1472.

Link to [NUREG 1556 Vol. 12 Rev. 1.](#)

Best Regards,

**Antonio J. Gonzalez, Health Physicist**

U.S. Nuclear Regulatory Commission, Region IV

Division of Radiological Safety & Security

Materials Licensing Branch

P: 817-200-1472

E: [Antonio.Gonzalez@nrc.gov](mailto:Antonio.Gonzalez@nrc.gov)