

From: [Jeremy Corwin](#)
To: [Roberto Torres](#)
Subject: [External_Sender] RE: NRC request for information - Central Montana Medical Center renewal application (25-18307-01)
Date: Monday, June 16, 2025 9:21:42 AM
Attachments: image001.png
Letter - Response to Questions - June 2025.pdf

Enclosed is a response letter. Let me know if you have any questions.
Thanks!

Jeremy L. Corwin, MHP, DABHP
Medical Physicist
Phone: (360) 736-6066 / Fax: (360) 736-7293
www.corwinhp.com



From: Roberto Torres <RobertoJ.Torres@nrc.gov>
Sent: Tuesday, June 10, 2025 9:39 AM
To: Jeremy Corwin <jeremy@corwinhp.com>
Subject: RE: NRC request for information - Central Montana Medical Center renewal application (25-18307-01)

Please provide the following commitment in a signed and dated letter (Item 8).

From: Roberto Torres
Sent: Tuesday, June 10, 2025 11:17 AM
To: jeremy@corwinhp.com
Subject: NRC request for information - Central Montana Medical Center renewal application (25-18307-01)

Mr. Corwin:

I am processing Central Montana Medical Center renewal application. Please provide the attached information identified in yellow highlight.

Thank you.

Roberto J. Torres, M.S.
Senior Health Physicist
U.S. Nuclear Regulatory Commission
Region IV
Division of Radiological Safety and Security
1600 East Lamar Boulevard
Arlington, TX 76011-4511

817-200-1189

June 16, 2025

Nuclear Regulatory Commission
Region IV
ATTN: Roberto Torres, M.S.
1600 E. Lamar Blvd
Arlington, TX 76011-4511

Re: Response to Questions - License Renewal 25-18307-01 (Central Montana Medical Center)

Dear Mr. Torres:

This letter is a response to emails received concerning the subject license renewal.

Regarding other RSO commitments, I am also an RSO on another NRC license 25-18361-01 (Community Medical Center) and WA State License WN-M048-1 (Virginia Mason). Regarding my time required for this license (Central Montana Medical Center), I can easily act as RSO as the use at this facility is only imaging, they receive unit doses, and they not require a significant amount of time. I conduct quarterly reviews of their program via phone and email and that is about 2-4 hours of time per quarter. My staff conducts an exhaustive annual Radiation Safety Program Review and ALARA Review through a personal visit to the facility, which I oversee and sign. I have a staff of nine additional physicists that are available to assist this site.

The In-House representative that will serve as the point of contact during my absence is the Radiology Manager, Heather Haynes (hhaynes@cmmccares.com).

Regarding ability to respond to questions and operational issues, as noted above we have a total of 10 physicists that are available to assist. The facility is able to call or email any questions and they get a response in a timely manner, usually immediately.

The maximum amount of time it will take for me to arrive at the facility is 10 hours. We feel this is reasonable considering that the facility is only using materials for imaging and there is very little risk of major incidents, especially considering that they receive unit doses. It is highly unlikely that there will be an emergency that requires my physical presence at this licensee.

Regarding the Training Program: We have developed and will implement and maintain written procedures for a program for training required under 10 CFR 19.12 for each group of workers including (i) topics covered, (ii) qualifications of the instructors (iii)

method of training, (iv) method for assessing the success of the training, (v) initial training, and (vi) annual refresher training.

Please contact me at (360) 736-6066 or at jeremy@corwinhp.com if there are any questions.

Sincerely,

A handwritten signature in black ink, appearing to read "J. Corwin". The signature is fluid and cursive, with a large initial "J" and a distinct "C".

Jeremy L. Corwin, MHP, CHP
Radiation Safety Officer
Central Montana Medical Center